

## Rajasthan Medical Services Corporation Limited

Gandhi Block, SwasthyaBhawan, Tilak Marg, C-Scheme, Jaipur-302005 (Raj.)

Phone No: 0141-2228066 ,2228064

Website: <http://rmsc.health.rajasthan.gov.in>

CIN:U24232RJ2011SGC035067

E\_mail :[cdprmsc@nic.in](mailto:cdprmsc@nic.in)

Ref. No.:- F.02(358)/RMSCL/Proc./Empanelment/NIPQ 01/2022/ 8761

Dated:- 14/12/2022

### Corrigendum – I

**Subject:- Amendment in Specifications and additions of items in bid.**

**Ref.:- NIB No. F.02(358)/RMSCL/Proc./Empanelment/NIPQ 01/2022/ 8721 Dated:09.12.2022**

The following amendments are made:

S. No	Existing Specification	Amended Specification																		
1	<p><b>LIST OF SPECIFICATION SURGICAL DRUG ITEMS</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 5%;">S. No.</th> <th style="width: 10%;">Code No.</th> <th style="width: 85%;">Name of approved item (s) with specification</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">1</td> <td style="text-align: center;">S-26</td> <td>                     Sterile Hypodermic Syringe with Needle attached, 24G, Single Use - 2 ml                     <ul style="list-style-type: none"> <li>Clear transparent chamber</li> <li>Prominent graduation</li> <li>Inert material gasket at the piston to minimize friction during movement &amp; prevent leakage and back flow</li> <li>Sharp needle ensuring minimum trauma during penetration</li> <li><b>As per IS 12050</b></li> <li>Packing: Needle should be attached with the syringe and packed in unit ribbon pack</li> <li>The words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container.</li> </ul> </td> </tr> <tr> <td style="text-align: center;">2</td> <td style="text-align: center;">S-27</td> <td>                     Sterile Hypodermic Syringe with Needle attached, 24G, Single Use - 5 ml                     <ul style="list-style-type: none"> <li>Clear transparent chamber</li> <li>Prominent graduation</li> <li>Inert material gasket at the piston to minimise friction during movement &amp; prevent leakage and back flow</li> <li>Sharp needle ensuring minimum trauma during penetration</li> <li><b>As per IS 12050</b></li> <li>Packing: Needle should be attached with the syringe and packed in unit ribbon pack</li> <li>The words "DESTROY AFTER</li> </ul> </td> </tr> </tbody> </table>	S. No.	Code No.	Name of approved item (s) with specification	1	S-26	Sterile Hypodermic Syringe with Needle attached, 24G, Single Use - 2 ml <ul style="list-style-type: none"> <li>Clear transparent chamber</li> <li>Prominent graduation</li> <li>Inert material gasket at the piston to minimize friction during movement &amp; prevent leakage and back flow</li> <li>Sharp needle ensuring minimum trauma during penetration</li> <li><b>As per IS 12050</b></li> <li>Packing: Needle should be attached with the syringe and packed in unit ribbon pack</li> <li>The words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container.</li> </ul>	2	S-27	Sterile Hypodermic Syringe with Needle attached, 24G, Single Use - 5 ml <ul style="list-style-type: none"> <li>Clear transparent chamber</li> <li>Prominent graduation</li> <li>Inert material gasket at the piston to minimise friction during movement &amp; prevent leakage and back flow</li> <li>Sharp needle ensuring minimum trauma during penetration</li> <li><b>As per IS 12050</b></li> <li>Packing: Needle should be attached with the syringe and packed in unit ribbon pack</li> <li>The words "DESTROY AFTER</li> </ul>	<p><b>LIST OF SPECIFICATION SURGICAL DRUG ITEMS</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 5%;">S. No.</th> <th style="width: 10%;">Code No.</th> <th style="width: 85%;">Name of approved item (s) with specification</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">1</td> <td style="text-align: center;">S-26</td> <td>                     Sterile Hypodermic Syringe with Needle/or with Needle attached , 24G, Single Use - 2 ml                     <ul style="list-style-type: none"> <li>Clear transparent chamber</li> <li>Prominent graduation</li> <li>Inert material gasket at the piston to minimize friction during movement &amp; prevent leakage and back flow</li> <li>Sharp needle ensuring minimum trauma during penetration</li> <li><b>Shall conform to relevant IS</b></li> <li>Packing: Needle should be attached with the syringe and packed in unit ribbon pack</li> <li>The words "DESTROY AFTER SINGLE USE" or equivalent should be writtn on Unit Container.</li> </ul> </td> </tr> <tr> <td style="text-align: center;">2</td> <td style="text-align: center;">S-27</td> <td>                     Sterile Hypodermic Syringe with Needle/or with Needle attached , 24G, Single Use - 5 ml                     <ul style="list-style-type: none"> <li>Clear transparent chamber</li> <li>Prominent graduation</li> <li>Inert material gasket at the piston to minimize friction during movement &amp; prevent leakage and back flow</li> <li>Sharp needle ensuring minimum trauma during penetration</li> <li><b>Shall conform to relevant IS</b></li> <li>Packing: Needle should be attached with the syringe and packed in unit ribbon pack</li> </ul> </td> </tr> </tbody> </table>	S. No.	Code No.	Name of approved item (s) with specification	1	S-26	Sterile Hypodermic Syringe with Needle/or with Needle attached , 24G, Single Use - 2 ml <ul style="list-style-type: none"> <li>Clear transparent chamber</li> <li>Prominent graduation</li> <li>Inert material gasket at the piston to minimize friction during movement &amp; prevent leakage and back flow</li> <li>Sharp needle ensuring minimum trauma during penetration</li> <li><b>Shall conform to relevant IS</b></li> <li>Packing: Needle should be attached with the syringe and packed in unit ribbon pack</li> <li>The words "DESTROY AFTER SINGLE USE" or equivalent should be writtn on Unit Container.</li> </ul>	2	S-27	Sterile Hypodermic Syringe with Needle/or with Needle attached , 24G, Single Use - 5 ml <ul style="list-style-type: none"> <li>Clear transparent chamber</li> <li>Prominent graduation</li> <li>Inert material gasket at the piston to minimize friction during movement &amp; prevent leakage and back flow</li> <li>Sharp needle ensuring minimum trauma during penetration</li> <li><b>Shall conform to relevant IS</b></li> <li>Packing: Needle should be attached with the syringe and packed in unit ribbon pack</li> </ul>
S. No.	Code No.	Name of approved item (s) with specification																		
1	S-26	Sterile Hypodermic Syringe with Needle attached, 24G, Single Use - 2 ml <ul style="list-style-type: none"> <li>Clear transparent chamber</li> <li>Prominent graduation</li> <li>Inert material gasket at the piston to minimize friction during movement &amp; prevent leakage and back flow</li> <li>Sharp needle ensuring minimum trauma during penetration</li> <li><b>As per IS 12050</b></li> <li>Packing: Needle should be attached with the syringe and packed in unit ribbon pack</li> <li>The words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container.</li> </ul>																		
2	S-27	Sterile Hypodermic Syringe with Needle attached, 24G, Single Use - 5 ml <ul style="list-style-type: none"> <li>Clear transparent chamber</li> <li>Prominent graduation</li> <li>Inert material gasket at the piston to minimise friction during movement &amp; prevent leakage and back flow</li> <li>Sharp needle ensuring minimum trauma during penetration</li> <li><b>As per IS 12050</b></li> <li>Packing: Needle should be attached with the syringe and packed in unit ribbon pack</li> <li>The words "DESTROY AFTER</li> </ul>																		
S. No.	Code No.	Name of approved item (s) with specification																		
1	S-26	Sterile Hypodermic Syringe with Needle/or with Needle attached , 24G, Single Use - 2 ml <ul style="list-style-type: none"> <li>Clear transparent chamber</li> <li>Prominent graduation</li> <li>Inert material gasket at the piston to minimize friction during movement &amp; prevent leakage and back flow</li> <li>Sharp needle ensuring minimum trauma during penetration</li> <li><b>Shall conform to relevant IS</b></li> <li>Packing: Needle should be attached with the syringe and packed in unit ribbon pack</li> <li>The words "DESTROY AFTER SINGLE USE" or equivalent should be writtn on Unit Container.</li> </ul>																		
2	S-27	Sterile Hypodermic Syringe with Needle/or with Needle attached , 24G, Single Use - 5 ml <ul style="list-style-type: none"> <li>Clear transparent chamber</li> <li>Prominent graduation</li> <li>Inert material gasket at the piston to minimize friction during movement &amp; prevent leakage and back flow</li> <li>Sharp needle ensuring minimum trauma during penetration</li> <li><b>Shall conform to relevant IS</b></li> <li>Packing: Needle should be attached with the syringe and packed in unit ribbon pack</li> </ul>																		



		SINGLE USE" or equivalent should be written on Unit Container.			<ul style="list-style-type: none"> <li>The words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container.</li> </ul>
3	S-28	Sterile Hypodermic Syringe with Needle attached, 22G, Single Use - 10 ml <ul style="list-style-type: none"> <li>Clear transparent chamber</li> <li>Prominent graduation</li> <li>Inert material gasket at the piston to minimise friction during movement &amp; prevent leakage and back flow</li> <li>Sharp needle ensuring minimum trauma during penetration</li> <li>As per IS 12050</li> <li>Packing: Needle should be attached with the syringe and packed in unit ribbon pack</li> <li>The words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container.</li> </ul>	3	S-28	Sterile Hypodermic Syringe with Needle/or with Needle attached, 22G, Single Use - 10 ml <ul style="list-style-type: none"> <li>Clear transparent chamber</li> <li>Prominent graduation</li> <li>Inert material gasket at the piston to minimize friction during movement &amp; prevent leakage and back flow</li> <li>Sharp needle ensuring minimum trauma during penetration</li> <li>Shall conform to relevant IS</li> <li>Packing: Needle should be attached with the syringe and packed in unit ribbon pack</li> <li>The words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container.</li> </ul>
4	S-29	Sterile Hypodermic Syringe with Needle attached, 22G, Single Use - 20 ml <ul style="list-style-type: none"> <li>Clear transparent chamber</li> <li>Prominent graduation</li> <li>Inert material gasket at the piston to minimise friction during movement &amp; prevent leakage and back flow</li> <li>Sharp needle ensuring minimum trauma during penetration</li> <li>z</li> <li>Packing: Needle should be attached with the syringe and packed in unit ribbon pack</li> <li>The words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container.</li> </ul>	4	S-29	Sterile Hypodermic Syringe with Needle/or with Needle attached, 22G, Single Use - 20 ml <ul style="list-style-type: none"> <li>Clear transparent chamber</li> <li>Prominent graduation</li> <li>Inert material gasket at the piston to minimize friction during movement &amp; prevent leakage and back flow</li> <li>Sharp needle ensuring minimum trauma during penetration</li> <li>Shall conform to relevant IS</li> <li>Packing: Needle should be attached with the syringe and packed in unit ribbon pack</li> <li>The words "DESTROY AFTER SINGLE USE" or equivalent should be written on Unit Container.</li> </ul>

**Addition of items:-**

S.N	Code No.	Name of approved item (s) with specification	Minimum labelled Shelf life (In months)	Packing Unit	Class As per Medical Device Rules 2017	Bid quantity
1	NRR-40	Braided polyester coated with Silicon 2-0 8X75CM 2XY-31 PLGT, BLUE & WHITE 16MM 1/2 Circle Taper cutting oval pledget	36 months	12 foils	Class C	3810
2	NRR-41	Braided polyester coated with Silicon 2-0 10X75 2XCV-305 PGT, BLUE & WHITE 25MM 1/2 Circle Taper Point oval pledget	36 months	12 foils	Class C	4950
3	NRR-42	Braided polyester coated with polybutylate 2-0 8X75CM 2X PLGT, BLUE & WHITE 16MM	36 months	12 foils	Class C	1246



		1/2 Circle Taper cutting oval pledget				
4	NRR-43	Braided polyester coated with polybutylate 2-0 10X75 2X PLGT, BLUE & WHITE 25MM 1/2 Circle Taper Point oval pledget	36 months	12 foils	Class C	2986
5	NRR-46	Braided polyester coated with Polybutylate 2, 26MM 1/2 Circle RC 75cm	36 months	12 foils	Class C	494
6	NRR-47	Braided polyester coated with Polybutylate 5, 55MM 1/2 Circle RC 75cm	36 months	12 foils	Class C	612
7	NRR-48	Monofilament polypropylene with PEG Additive 3-0 90CM 2XVF-20, BLUE 26MM 1/2 Circle Taper Point	36 months	12 foils	Class C	8260
8	NRR-49	Monofilament polypropylene with PEG Additive 2-0 90CM 2XV-20, BLUE 30MM 1/2 Circle Taper Point	36 months	12 foils	Class C	6366
9	NRR-50	Monofilament polypropylene with PEG Additive 4-0 90CM 2XCV-23, BLUE 17MM 1/2 Circle Taper Cut	36 months	12 foils	Class C	5864
10	NRR-51	Monofilament polypropylene with PEG Additive 5-0 90CM 2XCV-23, BLUE 17MM 1/2 Circle Taper Pont	36 months	12 foils	Class C	6669
11	NRR-52	Monofilament polypropylene with PEG Additive 6-0 75CM 2XCV-22, BLUE 13MM 1/2 Circle Taper Pont	36 months	12 foils	Class C	4941
12	NRR-53	Monofilament polypropylene with PEG Additive 7-0 60CM 2XKV-1, BLUE 9MM 3/8 Circle Taper cutting	36 months	12 foils	Class C	4432
13	NRR-77	Laprosopic Knotless Polydioxanone with Fixation Surgical Suture Self fixation device with autolock mechanism made up of Polydioxanone with Fixation Tab reverse cutting 36 mm & 45 cm Suture Size 1	36 months	12 foils	Class C	1802
14	NRR-78	Laprosopic Knotless Polydioxanone with Fixation Surgical Suture Self fixation device with autolock mechanism made up of Polydioxanone with Fixation Tab Taper point 36 mm & 45	36 months	12 foils	Class C	1575



		cm Suture Size 1-0				
15	NRR-89	Absorbable Surgical Suture (Synthetic )Coated Polyglactin/PGA 910 Voilet 6- 0 RC micro point ¼ circle 8 mm 45 cm '2670'	36 months	12 foils	Class C	2064

**Note:-**

- It may be noted that if any further amendments are issued then a corrigendum will be published and informed.
- Rest of the terms and conditions will remain the same.

  
 Executive Director (Proc.)  
 RMSCL