



Rajasthan Medical Services Corporation Limited

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur-302005 (Raj.)

Phone No: 0141-2228066 , 2228064

Website: <http://rmscl.health.rajasthan.gov.in>

CIN:U24232RJ2011SGC035067

E_mail : edprmsc@nic.in

Ref. No.: F.02(84)/RMSCL/Proc/Non Drug /NIB-16/2020/ 416 Dated: 30/10/2020

FINAL LIST OF BIDDERS DECLARED RESPONSIVE / NON RESPONSIVE

NIT Ref. No.:- F.02(84)/RMSCL/Proc/Non Drug /NIB-16/2020/347 Dated:09.10.2020

(TECHNICAL BID OPENED ON 14.10.2020)

| S. No. | Name of Bidder | Quoted Item No. | Bidder Status Mfg/Imp orter | Responsive Item Code No. | Non Responsive Item Code No. |
|--------|-------------------------------------|-----------------|-----------------------------|--|--|
| 1 | 2 | 3 | 4 | 5 | 6 |
| 1 | Amita projects | NE-S8 | Mfg | NE-S8 Although sample does not have Rajasthan Vidhansabha Picture in black & white color on left side and below to this picture Rajasthan Sarkar not written but your bid held responsive subject to condition to supply item as per RMSCL specifications | |
| 2 | Deeksha the clothing style, Jaipur | NE-S8 | Mfg | Nil | Non Responsive 1. Monthly supply commitment submitted in annexure VII is less than required commitment 40 lacs fabric mask 2. Annexure VIII of Manufacturing capacity and supply commitment not submitted. 3. GSM of outer layer 105.16.instead of 75-90. 4. GSM of inner layer is 72.10 instead of 95-110 |
| 3 | Emerge Glass India Pvt. Ltd. Jaipur | NE-S8 | Mfg | Nil | Non Responsive 1. GSM of inner layer is 80.92 instead of 95-110. |
| 4 | Jackson Care Product Jaipur | NE-S8 | Mfg | Nil | Non Responsive 1. GSM of outer layer 110.12.instead of 75-90. 2. GSM of inner layer 130.10 instead of 95-110. |
| 5 | Nandani Creation Ltd. Jaipur | NE-S8 | Mfg | Nil | Non Responsive 1. GSM of inner layer 126.00.instead of 95-110. |
| 6 | Paragon Apparel Pvt. Ltd. | NE-S8 | Mfg | NE-S8 | Nil |

09 ✓ →



| S. No. | Name of Bidder | Quoted Item No. | Bidder Status Mfg/Imp orter | Responsive Item Code No. | Non Responsive Item Code No. |
|--------|---------------------------------|-----------------|-----------------------------|--|--|
| 7 | Red Cherry Jaipur | NE-S8 | Mfg | Nil | Non Responsive 1. GSM of inner layer 76.12 instead of 95-110. |
| 8 | Rodas Impex(P) Ltd. Bhilwara | NE-S8 | Mfg | Nil | Non Responsive 1. GSM of middle layer is 64.16 instead of 40-60. 2. GSM of inner layer 129.16 instead of 95-110. |
| 9 | Shubham Creation Jaipur | NE-S8 | Mfg | Nil | Non Responsive 1. GSM of outer layer is 112.16 instead of 75.90. 2. GSM of middle layer is 68.00 instead of 40-60. 3. GSM of inner layer 70.19 instead of 95-110. |
| 10 | Tijaria Polypipes ltd. Jaipur | NE-S8 | Mfg (MSME Medium) | Nil | Non Responsive 1. GSM of inner layer 75.42 instead of 95-110. |
| 11 | Vevina Garments | NE-S8 | Mfg (MSME) small | NE-S8 Although length of elastic strings is 14.20 cm instead of 13.0± 0.2 cm but your bid held responsive subject to condition to supply item as per RMSCL specifications | |
| 12 | Vikas Jaipur | NE-S8 | Mfg (MSME micro) | NE-S8 Although length of mask is 15.50cm instead of 15.8 ± 0.2 but your bid held responsive subject to condition to supply item as per RMSCL specifications | |
| 13 | Yufta LLP Jaipur | NE-S8 | Mfg | NE-S8 | Nil |
| 14 | Mamta Industries | NE-S8 | Mfg | Nil | Non Responsive Blank bid submitted |
| 15 | MAUKTI K RETAIL PRIVATE LIMITED | NE-S8 | Mfg | Nil | Non Responsive Firm has not submitted bid security fees, tender fees, RISL Processing fees physically |

NOTE:- Online financial bid will be opened on 30.10.2020 at 01:30 PM

Handwritten signature/initials

Handwritten signature
(Shriniwas Meena)
Executive Director (Proc.)
RMSCL