

Rajasthan Medical Services Corporation Limited

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur-302005 (Raj.)

Phone No: 0141-2228066 , 2228064

Website: <http://rmsc.health.rajasthan.gov.in>

CIN:U24232RJ2011SGC035067

E_mail : edprmsc@nic.in

Ref. No.: F.02(86)/RMSCL/Proc/N-95 mask /NIB-19/2020/41

Dated: 26/2/2021

**FINAL LIST OF BIDDERS DECLARED RESPONSIVE /
NON RESPONSIVE**

NIT Ref. No.:- F.02(86)/RMSCL/Proc/N-95 mask /NIB-19/2020/442 Dated:01.12.2020

(TECHNICAL BID OPENED ON 29.12.2020)

S. No.	Name of Bidder	Quoted Item Code No.	Bidder Status Mfg / Importer	Responsive Bidders / Items code no.	Non responsive Bidders / Items code no.
1	2	3	4	5	6
1	Aamr LifeSciences Pvt. Ltd.	N-95	MSME Unit	Nil	Non responsive due to N-95 mask samples testing in laboratory does not meet the requirement as (Particulate filtration efficiency)
2	Accumax Lab devices Pvt. Ltd.	N-95	Mfg	Nil	Non responsive due to N-95 mask samples testing in laboratory does not meet the requirement as (Particulate filtration efficiency)
3	Amita Projects	N-95	MSME Unit	Nil	Non responsive due to N-95 mask samples testing in laboratory does not meet the requirement as (Particulate filtration efficiency)
4	Arihant Packwell	N-95	Mfg	Nil	Non responsive due to N-95 mask samples testing in laboratory does not meet the requirement as (Particulate filtration efficiency)
5	Deepali Designs & Exhibits Pvt. Ltd.	N-95	MSME Unit	Nil	Non responsive due to N-95 mask samples testing in laboratory does not meet the requirement as (Particulate filtration efficiency)
6	Devesh Fab Tex	N-95	Mfg	Nil	Non responsive due to N-95 mask samples testing in laboratory does not meet the requirement as (Particulate filtration efficiency and Breathing Resistance)
7	Emerge Glass India Pvt. LTd.	N-95	Mfg	N-95	Nil
8	H.Q Lamps Manufacturing Co. Pvt. Ltd.	N-95	Mfg	Nil	Non responsive due to N-95 mask samples testing in laboratory does not meet the requirement as (Particulate filtration efficiency)
9	Karam Industries	N-95	Mfg	N-95	Nil
10	Kingfa Science & Technology India Ltd.	N-95	Mfg	N-95	Nil
11	Lars Medicare Pvt. Ltd.	N-95	Mfg	N-95	Nil
12	Mahaveer	N-95	MSME	Nil	Non responsive due to N-95 mask

Handwritten signatures and initials at the bottom of the page.

मुख्यमंत्री निःशुल्क दवा योजना




S. No.	Name of Bidder	Quoted Item Code No.	Bidder Status Mfg / Importer	Responsive Bidders / Items code no.	Non responsive Bidders / Items code no.
	Polypack		Unit		samples testing in laboratory does not meet the requirement as (Particulate filtration efficiency and Breathing Resistance)
13	Megha International	N-95	Importer	Nil	Non responsive due to turnover submitted by firm does not meet the requirement of bid condition 2(b)
14	Nandani Creation Ltd.	N-95	MSME Unit	Nil	Non responsive due to N-95 mask samples testing in laboratory does not meet the requirement as (Particulate filtration efficiency)
15	P.A.R.K. Non woven Pvt. Ltd.	N-95	Mfg	N-95	Nil
16	Paramount Surgimed Ltd.	N-95	Mfg	Nil	Non responsive due to N-95 mask samples testing in laboratory does not meet the requirement as (Particulate filtration efficiency)
17	Rashmi Rare Earth Ltd.	N-95	Mfg	Nil	Non responsive due to N-95 mask samples testing in laboratory does not meet the requirement as (Particulate filtration efficiency)
18	Rodas Impex Pvt. Ltd.	N-95	MSME Unit	Nil	Non responsive due to N-95 mask samples testing in laboratory does not meet the requirement as (Particulate filtration efficiency)
19	Rosmerta Technologies Ltd.	N-95	Mfg	N-95	Nil
20	Samkit Textile Mills Pvt. Ltd.	N-95	Mfg	N-95	Nil
21	Sanmati Packaging	N-95	Mfg	N-95	Nil
22	Shasti Spuntec India Pvt. Ltd.	N-95	Mfg	Nil	Non responsive due to N-95 mask samples testing in laboratory does not meet the requirement as (Particulate filtration efficiency and Splash resistance)
23	Sky Lifecare Pvt. Ltd.	N-95	Mfg	Nil	Non responsive due to 1. Firm submitted BIS certificate and sample with the name of Prince enterprises. 2. Firm submitted dealership certificate but as per bid condition 2(a), only manufacturer is allowed to participate in the bid.
24	Suchi Industries Ltd.	N-95	Mfg	N-95	Nil
25	Tijaria Polypipes Ltd.	N-95	Mfg (MSME)	Nil	Non responsive due to N-95 mask samples testing in laboratory does not meet the requirement as (Particulate filtration efficiency)

मुख्यमंत्री निःशुल्क दवा योजना

S. No.	Name of Bidder	Quoted Item Code No.	Bidder Status Mfg / Importer	Responsive Bidders / Items code no.	Non responsive Bidders / Items code no.
26	UNICLAN HEALTHCARE PRIVATE LIMITED	N-95	Mfg	Nil	Non responsive due to N-95 mask samples testing in laboratory does not meet the requirement as (Particulate filtration efficiency)
27	Vijay Sabre Safety Pvt. Ltd.	N-95	Mfg	Nil	Non responsive due to N-95 mask samples testing in laboratory does not meet the requirement as (Particulate filtration efficiency)
28	Zkee International	N-95	Mfg	Nil	Non responsive due to N-95 mask samples testing in laboratory does not meet the requirement as (Particulate filtration efficiency)

Online financial bid will be opened on 26.02.2021 at 04:00 PM.


Executive Director (Proc.)
RMSCL

W

SIRY