



## Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

E-mail: cdepmrmse-rj@nic.in

CIN : U24232RJ2011SGC035067

Website: <http://rmse.health.rajasthan.gov.in>

No. F-8( ) RMSC/EPM/M-1/NIB-686/2022-23/ 1166

Dated: 20/3/23

All Principals Medical Colleges,  
All the Superintendents of Medical Colleges & Attached Hospitals.  
Director (PH/HA/IEC/AIDS/RCH)

**Sub:- Approval/Purchase of Item, 'Breast Pump Electric (Professional)- Hospital Use (Multi user)', for the Rate Contract and supply, ending on 28.02.2025.**

Please find enclosed herewith approved firm and Breast Pump Electric (Professional)- Hospital Use (Multi user), for the rate contract and supply ending on 28.02.2025. The rate contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website [www.rmse.health.rajasthan.gov.in](http://www.rmse.health.rajasthan.gov.in)

In reference to the above rate contract No. F-8( ) RMSC/EPM/M-1/2021-22/NIB-686/.....Dated..... the supplier has executed the agreement with RMSC, as enclosed herewith for your information & necessary action. This bears an approval of M.D., RMSC, Jaipur.

Executive Director (EPM)  
Rajasthan Medical Services Corporation Ltd, Jaipur

No. F-8( ) RMSC/EPM/M-1/NIB-686/2022-23/ 1166

Dated: 20/3/23

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSC.
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital- Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All BME, RMSC
21. M/s Medela India Pvt Ltd., Gurgaon.
22. Guard file.

Executive Director (EPM)  
Rajasthan Medical Services Corporation Ltd, Jaipur

# Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: http://rmisc.health.rajasthan.gov.in

No. F-8( ) RMSC/EPM/M-1/NIB-686/2022-23/166

Dated: 20/3/23

## Rate Contract and Price Details

S.N.	Item	Detail
1.	Name of Equipment / Item	Breast Pump Electric (Professional)- Hospital Use (Multi user)
2.	Name & Address of Firm	M/s Medela India Pvt Ltd. 602-603, 6 floor MD Megapolis, Sohna Road Sec-48, Gurgaon Email: uday.kamath@medela.com
3.	Bidder Status	Importer
4.	Performance Security	Rs. 604000.00 vide DD no. 216874/18.08.2020 (Axis Bank) & Rs. 125000.00 Vide DD No. 016940/06.03.2022 (Axis Bank) Total Rs. 729000.00
5.	Brand/ Make	Medela
	Model	Symphony
6.	Packing Unit	Each
7.	R/C No. & Validity	R.C. No. 795 (Valid Up to 28.02.2025)
8.	Indicative Quantity	200
9.	<b>Rates for supply to RMSC from Suppliers</b>	
	(i) New Basic Rate (in Rs. )	122786.00
	(ii) IGST @ 18% (in Rs. )	22101.48
	(iii) CGST (in Rs. )	0.00
	(iv) SGST (in Rs. )	0.00
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs. )	144887.48
10.	<b>Rate For Supply to Consignee From RMSC</b>	
	(i) Basic Rate (in Rs.)	122786.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	6139.30
	(iii) CGST@ 9% { On (i)+(ii)} (in Rs.)	11603.28
	(iv) SGST@ 9% { On (i)+(ii)} (in Rs.)	11603.28
	(v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	152131.86
11.	<b>TECHNICAL SPECIFICATION OF BREAST PUMP ELECTRIC (PROFESSIONAL)- HOSPITAL USE (MULTI USER)</b>	
	1. Technical characteristics (specific to this type of device)	1. Pumping frequency 30 to 80 CPM or more and user adjustable. 2. Cushion inserted inside the breast cup so that it does not hurt the mother. 3. Suction Pressure 100 to 250 mm hg; user adjustable. 4. Able to express milk from both breasts simultaneously. 5. Collection bottles can be used for storage of milk. 6. Double alternating pumps/double cycling pumps.

*Devin Raj*

		7. Should be motorized breast pump units. 8. Should be hospital grade.
<b>2. PHYSICAL CHARACTERISTICS</b>		
2.1	Dimensions (metric)	Portable
2.2	Weight (lbs, kg)	Compact unit (weight less than 4 kg)
2.3	Configuration	LCD/LED display suction timing
2.4	Noise (in dB)	<60db
2.5	Heat dissipation	
2.6	Mobility, portability	
2.7	Power Requirements	220-240 V AC <sub>i</sub> + 10%, 50-60Hz power supply; 5A plug; TYPE D
2.8	Tolerance (to variations, shutdowns)	± 10% of input AC.
2.9	Protection	1. Electrical protection by reset table over current breakers or replaceable fuses. 2. Back flow protection of milk.
<b>3. ACCESSORIES, SPARE PARTS, CONSUMABLES</b>		
3.1	Accessories, Spare parts, Consumables	1. Reusable collection bottles along-with breast cups - 10 sets. 2. All kinds of tubes - 12 sets (if applicable). 3. Breast pump Valve and Membrane (Pack of 4 Valves and 2 membranes) 25 No. 4. Following accessories shall be supplied free of cost- <ul style="list-style-type: none"> <li>• Polypropylene BPA free Milk container with screw cap of Size— 150/200 ml; Qty 50. It should have level marking in milliliter.</li> <li>• Borosilicate glass beaker with level marking in milliliter:- 500ml – qty 02nos. 2000ml- qty 02nos</li> </ul> 5. Other accessories required for optimum functioning of the equipment.
<b>4. ENVIRONMENTAL AND DEPARTMENTAL CONSIDERATIONS</b>		
4.1	Atmosphere/Ambiance (air conditioning, humidity, dust...)	1. Operating condition: Capable of operating continuously in ambient temperature of 10 to 40°C and relative humidity of 15 to 90% in ideal circumstances. 2. Storage condition: Capable of being stored continuously in ambient temperature of 0 to 50°C and relative humidity of 15 to 90%.
4.2	User's care, Cleaning, Disinfection and Sterility issues	1. Disinfection: Parts of the Device that are designed to come into contact with the patient or the operator should either be capable of easy disinfection or be protected by a single use/disposable cover.
<b>5. STANDARDS AND SAFETY</b>		
5.1	Certificates /Performance and safety standards	1. Should be European CE with notified body/FDA (US) approved product. 2. Manufacturer/supplier should have ISO 13485 certificate for quality standard.
<b>6. TRAINING AND INSTALLATION</b>		
6.1	Pre-installation requirements: nature, values, quality, tolerance	Supplier to perform installation, safety and operation checks before handover.
6.2	Requirements for sign-off	Certificate of calibration and inspection from the factory.
6.3	Training of staff (medical,	Training of users in operation and basic maintenance shall be

*[Signature]*

	paramedical, technicians)	provided.
<b>7. WARRANTY AND MAINTENANCE</b>		
7.1	Warranty	3 years
7.2	Service contract clauses, including prices	1. Warranty of three years with free servicing (min. 3 in a year) during warranty.
<b>8. DOCUMENTATION</b>		
8.1	Operating manuals, service manuals, other manuals	1. User and maintenance manuals to be supplied in Hindi/English. 2. Certificate of calibration and inspection to be provided. 3. List to be provided of equipment and procedures required for local calibration and routine maintenance. 4. List to be provided of important spares and accessories. 5. Contact details of manufacturer, supplier and local service agent to be provided.
<b>9. NOTES</b>		
9.1	Service Support Contact details	Contact details of manufacturer, supplier and local service agent to be provided.
9.2	Recommendations or warnings	Any warning signs would be adequately displayed.

- The agreement signed between RMSCL and the firm forms integral part of the rate contract.

Executive Director (EPM)  
RMSC, Jaipur  
Dated: 20/3/23

No. F-8() RMSC/EPM/M-1/NIB-686/2022-23/ 1186

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PII)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All BME, RMSC
21. M/s Medela India Pvt Ltd., Gurgaon.
22. Guard file.

Executive Director (EPM)  
RMSC, Jaipur