



Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

E-mail: edepmrmisc-rj@nic.in

CIN : U24232RJ2011SGC035067

Website: <http://rmisc.health.rajasthan.gov.in>

No. F-8() RMSC/EPM/M-1/NIB-667/RC-824/2023-24/ 701

Dated: 2/11/23

All Principals Medical Colleges,
All the Superintendents of Medical Colleges & Attached Hospitals.
Director (PH/HA/IEC/AIDS/RCH)

Sub:- Approval/Purchase of Item, 'Deep Freezer (-) 20 Degree', for the Rate Contract and supply, ending on 31.10.2025.

Please find enclosed herewith approved firm and 'Deep Freezer (-) 20 Degree', for the rate contract and supply ending on 31.10.2025. The rate contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website www.rmisc.health.rajasthan.gov.in

In reference to the above rate contract No. F-8() RMSC/EPM/M-1/NIB-667/RC-824/2023-24/Dated..... the supplier has executed the agreement with RMSC, as enclosed herewith for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur

Executive Director (EPM)
RMSCL, Jaipur
Dated: 2/11/23

No. F-8() RMSC/EPM/M-1/NIB-667/RC-824/2023-24/ 701

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital- Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW) (RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All BME, RMSC
21. M/s Authentic Instruments, Jaipur.
22. Guard file.

Executive Director (EPM)
RMSCL, Jaipur



Rajasthan Medical Services Corporation Limited, Jaipur
Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065
CIN:U24232RJ2011ISGC035067

E-mail : edepmrmsc-rj@nic.in
Website: <http://rmsc.health.rajasthan.gov.in>

No. F-S() RMSC/EPM/M-1/NIB-667/RC-824/2023-24/ 701

Dated: 21/11/23

Rate Contract and Price Details

S.N.	Item	Detail
1.	Name of Equipment / Item	Deep Freezer (-) 20 Degree
2.	Name & Address of Firm	M/s Authentic Instruments, 31 RIICO Indl. Area Jhotwara, Kamani Choraha, Jaipur Email: support@authenticinstruments.com , calibration@authenticinstruments.com
3.	Bidder Status	Manufacturer
4.	Performance Security	Rs. 65150.00 vide BG No. 2920IGP003091823 dated 17.10.2023 valid up to 31.03.2029 (Bank of Baroda)
5.	Brand/ Make	Authentic
	Model	X14.3DF-300
6.	Packing Unit	Each
7.	R/C No. & Validity	R.C. No. 824 (Valid Up to 31.10.2025)
8.	Indicative Quantity	30
9.	Rates for supply to RMSC from Suppliers	
	(i) New Basic Rate (in Rs.)	184000.00
	(ii) IGST (in Rs.)	0.00
	(iii) CGST @ 9% (in Rs.)	16560.00
	(iv) SGST @ 9% (in Rs.)	16560.00
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	217120.00
10.	Rates For Supply to Consignee From RMSC	
	(i) Basic Rate (in Rs.)	184000.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	9200.00
	(iii) CGST@ 9% { On (i)+(ii)} (in Rs.)	17388.00
	(iv) SGST@ 9% { On (i)+(ii)} (in Rs.)	17388.00
	(v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	227976.00
11.	Technical Specifications for Deep Freezer (-20⁰):- 1. The system should be vertical standing model. 2. The system should have 7 segment/LCD electronic control unit with digital temperature display on panel. 3. The outer body of system is made of powder coated corrosion –resistant sheet and inner body high grade stainless steel/ABS plastic. 4. The system should have volume capacity of 350 liters ± 10 liters	



5. The temperature range should be (-20⁰ C to -25⁰ C) or below.
6. The system should be microprocessor controlled.
7. The system should have 24-hour backup for Uninterrupted 24 X 7 electronic temperature circular chart recorder / integrated digital temperature cum controller with inbuilt 7 days inkless circular chart recorder with supply of 200 free charts and should be provision to connect with central (temperature) monitoring system/data logger for monitoring the temperature and should have capacity to store more than 5 year data and data can be transfer by USB.
8. The system should have CFC Free refrigerant.
9. The entire cabinet should be able to maintain uniform temperature in each drawer.
10. The system should have PUF insulation.
11. The system should have minimum 4 or more in number drawer/shelves with single/individual lids.
12. The system should have a built- in audible & visible alarm, if temperature fluctuates from its set-point range & door open warning system.
13. Firm should bring calibrated digital temp. Monitor/Thermometer (NABL Accredited) to check entire cabinet temp at the time of demonstration.
14. Should be compliant as per applicable standard with European CE (Notified body) or USFDA (Valid documentation should be submitted in technical bid).
15. Manufacturer should be compliant with ISO 13485 accredited from IAF (Valid documentation should be submitted in technical bid).
16. A servo stabilizer inbuilt/external to be supplied with the system
17. The system should be operable at 220-230 V & 50-60 Hz
18. Supply with four wheel base for easy operation with locking mechanism.
19. Guarantee: Three years on equipment.

- The agreement signed between RMSCL and the firm forms integral part of the rate contract.

Executive Director (EPM)
RMSCL, Jaipur

No. F-8() RMSC/EPM/M-1/NIB-667/RC-824/2023-24/ 701

Dated: 2/11/23

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/(RH)/(Hospital Admn.)



15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All BME, RMSC
21. **M/s Authentic Instruments, Jaipur.**
22. Guard file.

[Handwritten signature]
22/1/13

Executive Director (EPM)
RMSCL, Jaipur