

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax: 0141-2228065

E-mail: edepmrmsc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: http://rmsc.health.rajasthan.gov.in

F-8()RMSC/EPM/M-4/NIB-684/2023-24/8/42

Dated 11 12/23

Rate Contract and Price Details

S.N.	Item		Detail	
1.	Name of Equipment / Item	Pulse Oxymeter (Finger	Pulse Oxymeter (Finger Tip).	
2.	Name & Address of Firm	BPL Medical Technologies P A-654 Ground Floor, Royal I Station, Malviya Nagar-302017 Email: sales.healthcare@bpl	Paradise Opp. Malviya Nagar Police	
3.	Bidder Status	Manufacturer	2	
4.	Performance Security	Performance Security Deposit through Bank Gurantee No. 009GT02233360002 Dated 02-12-2023 Rs. 33,07,800 Valid Up to 31.05.2028		
5.	Brand/Model	BPL Smart Oxy		
	IISN Code	90181990		
6.	Packing Unit	. Each		
7.	R/C No. & Validity	R.C. No 825 Valid Up to 3	0.11.2025)	
8.	Indicative Quantity	57000		
9.	Rates for supply to RMSC from Suppliers			
	(i) Basic Rate (in Rs.)		1035.00	
	(ii) IGST @ 12% (in Rs.)		124.20	
	(iii) CGST (in Rs.)		0	
	(iv) SGST(in Rs.)		0.00	
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)		1159.20	
10.	Rate For Supply to Consignee From RMSC			
	(i) Basic Rate (in Rs.)		1035.00	
	(ii) RMSC Surcharge@5%	On (i) (in Rs.)	51.75	
	(iii) CGST@ 6% { On (i)+(ii)} (in Rs.)		65.20	
	(iv) SGST@ 6% { On (i)+(ii)} (in Rs.)		65.20	
	(v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)		1217.16	
11.		ulse Oximeter (Fingertip)		
	 Technical specifications of Pulse Oximeter (Fingertip) This fingertip oxymeter is intended for measuring the pulse rate and functional oxygen saturation (SpOtherough a patient's finger from adult to pediatric patient. It should display SpO₂, PR, Plethysmogram and Perfusion Index. Best Features: Auto Power Off, Water Resistant, Buzzer- Visual Alarm, Dual color OLED display/seve segment LED. Low battery indication. Quick, Accurate and Easy To Use - It's newly designed monitor gives instant readings of blood oxyge saturation (SpO₂), pulse rate (PR), Plethysmogram and displays them on a large screen. Wavelengt display clearly shows inconsistencies. Its display can rotate in four directions. The quoted model should be European CE (Notified Body)/USFDA/BIS Certified. 			

of



Type of Product	Digital Fingertip Pulse Oxymeter		
SpO ₂ Measurement Range	70% - 100%		
Pulse Measurement Range	30 - 250bpm		
SpO ₂ Measurement Accuracy	70% - 100% ± 2 %		
Pulse Measurement Accuracy	30 - 99bpm ± 2bpm		
	100-250bpm ± 2%		
Pulse Indication	0.2 - 1.0% ±0.2digits;		
	1.1-20.0% ±20%		
Battery Type	2*AA/AAA 1.5V Alkaline Batteries Type	V	
Warranty	2 Year		

- (a) Bidders are requested to send with bid, printed descriptive literature (catalogue) of the quoted goods.
- (b) If bidder had supplied or had a rate contract of quoted goods with any other Govt. institution within one year, he may be asked to provide copies of purchase orders, invoices and rate contracts.
- The agreement signed between RMSCL and the firm forms integral part of the rate contract

Executive Director (FPM)
RMSC, Japur
and

F-8()RMSC/EPM/M-4/NIB-684/2023-24/ 842

Copy forwarded to the following for information & necessary action:

- 1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
- 2. P.S. to ACS/Principal Secy/Secy. Medical Health & FW Deptt. Rajasthan, Jaipur.
- 3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
- 4. P.S. to Mission Director NHM Jaipur.
- 5. PA to Managing Director, RMSCL
- 6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
- 7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar& Udaipur.
- 8. Principal R.U.H.S. College of Dental Science Jaipur.
- 9. Principal, RUHS College of Medical Scinces, Jaipur.
- 10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Scinces, Jaipur).
- 11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
- 12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital-Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
- 13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
- 14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
- 15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
- 16. All P.M.O.'s.
- 17. All CM&HO's.
- 18. All Dy. CM&HO's.
- 19. AGM (IT), RMSC, Jaipur
- 20. All Zonal BME are directed to carry out inspection of the supplied equipment as per RMSCL Letter no. 2513 dated 10-07-2018.
- 21. BPL Medical Technologies Private LimitedA-654 Ground Floor, Royal Paradise Opp. Malviya Nagar Police Station, Malviya Nagar-302017Email: sales.healthcare@bpl.in

22. Guard file.

Executive Director (HPM)
RMSC, Jaipur