

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: http://rmisc.health.rajasthan.gov.in

No. F-8() RMSC/EPM/M-1/NIB-698/2022-23/40

Dated: 19/4/23

Rate Contract and Price Details

S.N.	Item	Detail
1.	Name of Equipment / Item	Indigenous Digital Audiometer (Pure-Tone Audiometer)
2.	Name & Address of Firm	M/s Labat Asia Pvt Ltd. Plot no. 1408, JPL Industrial Area, Sector-82, Mohali-160062, Email: info@labatasia.com
3.	Bidder Status	Manufacturer
4.	Performance Security	ICICI Bank BG no. 3416NDDG00000824 dated 12.04.2023 valid up to 30.09.2027 Rs. 276000.00
5.	Brand/ Make	Labat Asia Pvt
	Model	Audiolite V4.0
6.	Packing Unit	Each
7.	R/C No. & Validity	R.C. No. 799 (Valid Up to 31.03.2025)
8.	Indicative Quantity	60
9.	Rates for supply to RMSC from Suppliers	
	(i) New Basic Rate (in Rs.)	175000.00
	(ii) IGST @ 5% (in Rs.)	8750.00
	(iii) CGST (in Rs.)	0.00
	(iv) SGST (in Rs.)	0.00
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	183750.00
10.	Rate For Supply to Consignee From RMSC	
	(i) Basic Rate (in Rs.)	175000.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	8750.00
	(iii) CGST@ 2.5% { On (i)+(ii)} (in Rs.)	4593.75
	(iv) SGST@ 2.5% { On (i)+(ii)} (in Rs.)	4593.75
	(v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	192937.50
11.	Technical Specifications of Indigenous Digital Audiometer (Pure-Tone Audiometer) -	
	<ul style="list-style-type: none"> • Essential • Power supply: 220 Volts, 50Hz • Transducer: headset with earphones (TDH 39) & supraaural ear cushions (M X41 AR), bone vibrator (Radio ear b-71), Talk over microphone. • Patch cords. • Stimuli Type: Pure tones (continuous, pulse tone, interuptor) & Warble tones (continuous, pulse tone, interuptor), Narrowband noise, speech noise, speech (monitored live voice and recorded voice) • Frequencies: 250Hz- to 8000Hz in octaves for air conduction 250Hz to 4000Hz in octaves for bone conduction • Intensity: <ul style="list-style-type: none"> ➤ 10dB HL to 120dB HL in steps of 5dB for air conduction ➤ 10dB HL to 70 dB HL in steps of 5 dB for bone conduction from 250Hz to 4000Hz or more. 	

Rf



- Power supply: UPS/Battery backup & mains.
- Self-calibration
- Data saving facility should be available, should be supplied with Laptop (Intel processor i3, 4gb RAM, 1 TB Hard Disk with latest window) with color Ink Tank Printer.
- Two years guarantee to be given with the equipment.
- Minimum of two routine maintenance services/inspections to be provided annually.
- CMC will be given @5% of net rate (Exclusive GST as applicable) & yearly escalation of 5% on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion guarantee period of two year.
- The Manufacture should have ISO 13485.
- Should have safety certificate from a competent authority or valid detail electrical and functional safety test. Copy of the certificate/test report shall be produced along with the technical bid.
- Final technical approval after demonstration.
- Price quoted should be inclusive of their installation at site.

- The agreement signed between RMSC and the firm forms integral part of the rate contract.

Executive Director (EPM)

RMSC, Jaipur

Dated: 19/4/23

No. F-8() RMSC/EPM/M-1/NIB-698/2022-23/40

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS, Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSC.
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/(RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All BME, RMSC
21. M/s Labat Asia Pvt Ltd, Mohali
22. Guard file.

Executive Director (EPM)

RMSC, Jaipur