

RAJASTHAN MEDICAL SERVICES CORPORATION LTD.
D- Block, SwasthyaBhawan, TilakMarg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC 035067

E-Mail - edepmrmsc-rj@nic.in

No. F-8() RMSC/EPM/M-3/NIB-611/2021-22/ 1062

Dated: 07.09.2021

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Amended technical specifications and bid submission date of bid document of item Pulse Oxymeter (Table Top) under NIB No. F-8() RMSC/EPM/M-3/21-22/NIB-611/948 dated 05.08.2021

In reference to subject cited above NIB-611, various representation received from the firms and issues raised by the bidder are examined by the competent authorities and technical committee. The following clarification/corrigendum/addendum is issued for inclusion in bid document & Technical Specification of items as below:-

Revised Technical Specifications of Pulse Oxymeter (Table Top)

1. Suitable for all types of patients- adult, pediatric, neonatal.
2. Compact table top pulse oximeter with colour monitor at least **3.5" TFT screen or more.**
3. SpO2 measurement range at least 40-70 and 70 to 99%, minimum gradation 1%.
4. Accuracy of SpO2 better than $\pm 1\%$ from range 40-70 and better than $\pm 3\%$ for range 70-99.
5. Pulse rate range at least 30 to 240bpm, minimum gradation 1bpm.
6. Accuracy of pulse rate better than ± 5 bpm.
7. Audiovisual alarms required: High and low SpO2 and pulse rate (operator variable settings), sensor disconnected, sensor failure, low battery.
8. Plethysmograph (may be in form of bar) display is mandatory.
9. Should have minimum 24 hrs trend memory for SpO2 & PR.
10. Display shows SpO2 (%), HR (bpm) and signal strength bar.
11. SpO2 technology should be **Nellcor/Masimo** and also sense in hypotension, shivering and motion.
12. Large display readable from distance, display covers of durable plastic.
13. User preset of high/low alarms on SpO2 and pulse rate monitoring.
14. Silencing feature for audio alarm.
15. Carrying handle should be providing for easy mobility.
16. Display reports system errors, probe failure, probe connection and built-in battery status.
17. Automatic switch from mains to batteries in case of power failure.
18. Power requirements: 220V, 50 Hz and internal re-chargeable battery (at least 4 hrs back up, automatic recharge)
19. Manufacturer should have ISO 13485 Certificate.
20. Quoted model should be European CE (Notified Body)/USFDA approved.
21. Guarantee: Three years on equipment from the date of installation.
22. CMC: CMC will be given @ 5 % (of net rate - exclusive of GST) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years.

Handwritten signature/initials

23. Table Top Pulse Oximeter should be supplied with standard set of:
 - a. Reusable SpO₂ sensors for Adult including connection cable.
 - b. Reusable SpO₂ sensors for Pediatrics, Clip on type including connection cable.
 - c. Reusable SpO₂ sensors for neonate, Wrap around type.
 - d. Forehead sensor.
24. Installation will be done by supplier free of cost.
25. Demonstration of equipment is must for technical approval.
26. The company should mention the make & model name/number of the quoted equipment and submit the technical brochure of the quoted model in the technical bid along with compliance sheet as per technical specifications.

Other terms and condition to be covered during guarantee period by the supplier:-

1. Response time- < 48 Hours after logged the complain in complain portal.
2. Service hours- As per health facility schedule.
3. Preventive Maintenance Schedule (PM)-As per OEM (Mention PM Schedule)
4. Calibration Schedule – As per OEM (Mention Schedule)
5. Up time- 95% (346 Days)
6. Technical & Application Support, Demonstrations & Trainings - As & when required
7. Toll Free No. (24*7)
8. Life of the equipment- As per OEM
9. Service engineer detail in Rajasthan

Revised bid schedule:-

E-bids are invited as per following revised time schedule:-

Existing Dates			Extended Dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
1	2	3	4	5	6
08.09.2021 11:00 a.m.	08.09.2021 06:00 p.m.	09.09.2021 03:00 p.m.	14.09.2021 11:00 a.m.	14.09.2021 06:00 p.m.	15.09.2021 03:00 p.m.

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

This bears an approval of Managing Director, Rajasthan Medical Services Corporation Limited, Jaipur.


Executive Director (EPM)
RMSCL, Jaipur