



CLARIFICATION/CORRIGENDUM/ADDENDUM

**Subject: Amendment in technical specification and processing fees of bid document of item (1)Air Purifier System (2) Calibrated Obstetric Under Buttock Drape (Labour Brass V-Drap), under No. F-80) RMSC/EPM/M-1/NIB-732/2022-23/934 date: 27.01.2023**

In reference to subject cited above NIB-732, various representation received from the firms and issues raised by the bidder are examined by the competent authorities and technical committee. The following clarification/corrigendum/addendum is issued for inclusion in bid document of items as below:-

**Revised Technical Specification Air Purifier System:-**

1. It shall have 7 stage air purification technologies that include:
  - 1) Pre-Filter washable
  - 2) Anti Bacterial (Easily Replaceable)
  - 3) HEPA Filter (Easily Replaceable)
  - 4) Activated Carbon Filter (Easily Replaceable)
  - 5) Tio2 Photo Catalyst Filter (Easily Replaceable)
  - 6) Ultra Violet Lamp-2 Nos.
  - 7) Ionizer
2. Airflow : 225-250 m3/hr
3. HEPA Filter/  
ULPA Filter : 0.3 micron @97%, 1 micron @99.97%
4. Voltage : 220 V
5. Power  
Consumption : 40-45 Watt
6. Min Noise Level : 35 db (± 3db)
7. Max Noise Level : 45 db (± 3db)
8. Timers : 1 to 12 Hours & Manual or IOT
9. Weight : 4.5-5.5 kg
10. Remote : Yes
11. Size : 330x180x50mm (±10 mm for L, B & H)
12. Filters need to be easily replaceable
13. COMPANY SHOULD BE ISO 9001:2015
14. Guarantee 1 Year.

**Revised Technical Specification of Calibrated Obstetric Under Buttock Drape (Labour Brass V-Drap)**

1. The obstetric drape graduated at 50mL measurements of postpartum
2. A pouch that holds more than 2000 ml of fluid.

 



3. Transparent (without color) polythene sheet (polythene sheet made of food grade plastic with minimum 120 micron. (Test report of food grade and micron from NABL LAB/GOVT APPROVED LAB to be submitted along with Technical Bid).
4. Manufacturer should have ISO: 22610 Certificate for Calibrated obstetric Under Buttock Drape. Valid document should be submitted in technical bid.
5. Samples (Qty-5 Nos.) to be submitted along with technical bid.
6. The Calibrated obstetric Under Buttock Drape shall be supplied with packing of low density polythene (LDPE) with liner low density polythene (LLDPE) octane based with minimum 120 micron. (Test report of food grade and micron from NABL LAB /GOVT APPROVED LAB to be submitted along with Technical Bid)
7. Firm should arrange Pre-dispatch inspection at ware house/factory

**Revised Tender Processing Fees**

Existing RISL Processing Fees	Revised RISL Processing Fees
1000.00	2000.00

**Note:-** Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

**Executive Director (EPM)  
RMSCL, Jaipur**