

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmse-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: http://rmse.health.rajasthan.gov.in

No. F-8() RMSC/EPM/M-3/22-23/NIB-743/ 1123

Dated: 10/3/23

CLARIFICATION/CORRIGENDUM/ADDENDUM

Sub:- Revised technical specifications and bid submission date of the rate contract for 3 Channel ECG Machine for under NIB No. F-8() RMSC/EPM/M-3/NIB-/2022-23/743/1008 dated 16-2-2023

In Reference to subject cited above and NIB-743, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in Bid document & Technical Specification of items as below:-

Revised Technical Specifications of 3 Channel ECG Machine

S.N.	Specification
1.	3 Channel ECG machine with colour TFT/LCD display. Size 5" or more.
2.	Automatic, Manual Recording Modes.
3.	Soft touch keyboard/ Qwerty keypad for data entry.
4.	Paper size: 60 mm or more.
5.	Should have visual alarm for open lead.
6.	Mains and Rechargeable battery operated.
7.	Inbuilt Rechargeable battery for at least 4 hours.
8.	Portable, lightweight
9.	Measurement facility with interpretation.
10.	Memory for 100 ECG or more.
11.	Automatic Lead Marking. Thermal printing.
12.	Recording Sensitivity: 2.5, 5, 10, 20 mm/mV.
13.	Defibrillation Protection: Internal
14.	Frequency Response of 0.05-150 Hz.
15.	Filters: Mains and Muscle interference.
16.	It should have USB Connection.
17.	It should have pacemaker detection.
18.	Safety standards Class II type CF test report shall be submitted in technical bid.
19.	Quoted model should be European CE (notified body) or USFDA approved.
20.	Guarantee 3 years on equipment from the date of Installation & minimum 06 months on Consumables.
21.	CMC shall be given 5% of net rate (Exclusive GST as applicable) plus service tax as applicable and escalation of 5 on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.
22.	Equipment should be supplied with compatible standard paper rolls (quantity 10 nos), additional Clamp Electrodes (1 set), Chest Electrodes (1 set), ECG Jelly (250 mlX2 bottles), ECG Cables (2 set), Power Cable (1 vet) Earth Cable-1, Guarantec Card & Operating Manual.
23.	The machine should be supplied with a good quality carry bag to carry ECG Machine
24.	Detail of Service Network of Firm in Rajasthan.

mf

S.N.	Specification
25.	The Company should mention the model name/number and submit the technical brochure of the quoted model in the Technical Bid.
26.	Paper Speed 12.5/25/50/100 mm/sec.

Revised Bid Schedule:-

E-bids are invited as per following revised time schedule:-

Existing Dates				Extended Dates			
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Last Date of submission of bid document fees	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Last Date of submission of bid document fees	Date of Opening of Technical Bid
1.	2.	3.	4.	5.	6.	7.	8.
<u>14.03.2023</u> <u>11:00 a.m.</u>	<u>14.03.2023</u> <u>6:00 p.m.</u>	<u>15.03.2023</u> <u>03:00 p.m.</u>	<u>15.03.2023</u> <u>03:00 p.m.</u>	<u>21.03.2023</u> <u>11:00 a.m.</u>	<u>21.03.2023</u> <u>6:00 p.m.</u>	<u>22.03.2023</u> <u>03:00 p.m.</u>	<u>22.03.2023</u> <u>03:00 p.m.</u>

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.


 Executive Director (EPM)
 RMSCL, Jaipur