

RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail -edepmrmisc-ri@nic.in

No. F-8()RMSC/EPM/M-7/NIB-530/ 2020-21/1146

Dated: 05/10/2021

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Amended Technical Specification of item Radiology High Frequency X-Ray Machine (100 Ma, 300 Ma, 500 Ma And 100MA Mobile) under No No. F-8() RMSC/EPM/M-7/NIB-530/ 2020-21/119 date: 15.10.2020.

Amended Technical Specification

S.N.	Item Name	Existing specification	Amended specification
1	High frequency X ray Machine (500 mA)	Tube focal spot selection switch	Tube focal spot selection switch/automatic selection.
2	High frequency X ray Machine (300 mA)	Tube focal spot selection switch	Tube focal spot selection switch/automatic selection.

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.


Executive Director (EPM)
RMSCL, Jaipur