

## Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur

Phone No.-2223887, Fax no.-0141-2228065

E-Mail: [edepmrmsc-ri@nic.in](mailto:edepmrmsc-ri@nic.in)

CIN:U24232RJ2011SGC035067

Website: [www.rmssc.health.rajasthan.gov.in](http://www.rmssc.health.rajasthan.gov.in)

F-80) RMSC/EPM/M-4/NIB-643/2021 117/6

Date:- 23/02/2022

### Clarification/Corrigendum/Addendum

**Subject: Amendment in technical specification, revised bid security detail of bid document of item LED PHOTOTHERAPY UNIT, under No. F-80) RMSC/EPM/M-4/NIB-643/2021-22/1340 Dated 22-11-21.**

#### **(A) Revised Technical Specifications of LED PHOTOTHERAPY UNIT**


1. **Life of LED shall be lifetime.**
2. Spectral Irradiance from 30 to 60  $\mu\text{W}\cdot\text{cm}^{-2}\cdot\text{nm}^{-1}$  at 45 cm distance between bed and light
3. Should have multilevel intensity control from minimum to maximum .
4. At the tilted position, the irradiance is 30  $\mu\text{W}\cdot\text{cm}^{-2}\cdot\text{nm}^{-1}$  at 45 cm distance between bed and light
5. Wavelength is 450 – 480 nm, and it's free from UV and IR radiation.
6. Effective surface area - 200 X 400 mm within a irradiance ratio of 0.4 (min/max irradiance)
7. Digital (LCD) Timer for monitoring therapy hours (resettable) & lamp usage hours (non resettable)
8. Have feather touch key pad
9. Cooling Fan/ heat sink is provided to dissipate the heat created by LED's
10. Light head should be compact to use along with the Radiant warmer & provided with tilting facility so that the unit is not coming directly under warmer.
11. Smooth Height adjustment mechanism & seamless height adjustment from minimum to maximum height
12. Minimum height – 1150 mm or less from the floor to use near the mother bed
13. Maximum height – 1500 mm or more from the floor to use with the incubator and Radiant warmer.
14. Coating: Epoxy/powder coated body for scratch and rust prevention and PU (Poly Urethane)/ABS/Fiber coating for plastic/ **Plastic with natural color (injection moulded with natural color).**
15. Mobility: Three/four castors; two rear castors provided with brakes
16. The base of the unit is such that it will go beneath any Incubator/bed/trolley, with minimum of 100 mm floor clearance
17. Should be supplied with white LED for patient examination
18. The unit should be supplied with reusable eye masks, Quantity: 10
19. Equipment should be US FDA approved / CE certified from notified body registered in European commission/BIS.
20. Warranty: Three years on equipment from the date of installation.

**(B) Revised bid security detail**

S N	Description of Goods under Procurement	Indicative Quantity (In Nos.)	Bid Value (In Rs.)	Minimum average Gross annual turnover for last three financial years (In Rs.)	Minimum average gross annual turnover for last three financial years for (MSMEs of Rajasthan) (In Rs.)	Bid Security (In Rs.)	Bid Security for MSME Units of Rajasthan (In Rs.)
1	LED PHOTOTHERAPY UNIT	500	1,35,00,000	81,00,000	40,50,000	270000	67500

**Note:** BF-3 Bid security Declaration Affidavit not applicable from 01-01-2022

**Note:-** Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

  
Executive Director (EPM)  
RMSCL, Jaipur