



**CLARIFICATION/CORRIGENDUM/ADDENDUM**

**Subject: Amended Specification and Time extension of item Multi Para Monitor of NIB No.F-8( ) RMSC/EPM/M-6/ NIB-633/2021-22 /1138 dated 30.09.2021**

**Revised Technical Specifications of Multi Para Monitor**

- 1) It should having ECG, SpO2, NIBP, Respiration and two Temperature as standard parameters.
- 2) Should have facility to display 7 lead ECG, RR, HR, SpO2, NIBP, and dual Temperature.
- 3) Should display 6 or more waveforms of selected parameters simultaneously, should have colour coding for different waveforms.
- 4) Should have inbuilt battery backup of at least three hours.
- 5) Display: Color LED/LCD touch screen display size of 12" or more.
- 6) Monitor should have both touch screen and knob to operate.
- 7) Should have dual temperature monitoring either in 0Celsius or 0Fahrenheit.
- 8) Should have facility for displaying multi - screen configurations.
- 9) Should be able to store & display at least 72 hours of graphical trends of all parameters.
- 10) Should be suitable for monitoring adult & pediatric & neonate patients.
- 11) The SpO2 technology should be Masimo/Nellcor/TruSignal and it should also sense in hypotension, shivering & motion.
- 12) Should have oscillometric technology for measurement of NIBP with Auto, START and Manual modes.
- 13) Should have different patient type selection.
- 14) The pulse rate should be displayed either with ECG or SpO2.
- 15) The respiration rate should be calculated through Impedance method.
- 16) Should be able to analyze arrhythmias & ST segment changes.
- 17) Memory should not wipe off when the power is turned off.
- 18) Should be able to give visual & audible alarms with three levels of volume adjustment.
- 19) Should have facility of connecting to the central monitoring station having facility for real time monitoring through Ethernet cable or Wireless Connectivity.
- 20) Wall mount stand should be provided free of the cost with the monitor.
- 21) Quoted model should be European CE (Notified Body)/USFDA approved.
- 22) Warrantee-Three years on equipment from the date of installation.
- 23) CMC: CMC will be given @ 5 % of Net Rate exclusive of GST (As applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Warrantee period of three years.
- 24) Equipment should be supplied with compatible 5 lead ECG cable (1 Nos.), IBP Cable (2 Nos.), SpO2 (Adult,Pediatric,Neonate)probe with extension cable (Each 1No.), NIBP Cuff with extension cable of Large Adult, Adult & Pediatric (2 each), Temperature probe (2 Nos.) & Operating / user Manual (1 No.).

15/11/21

## **B-Revised bid schedule:-**

- E-bids are invited as per following revised time schedule:-

<b>Existing Dates</b>			<b>Extended Dates</b>		
<b>Last date &amp; time for sale of bid form</b>	<b>Last date &amp; time of receipt of bid form</b>	<b>Date &amp; time of opening of technical bid</b>	<b>Last date &amp; time for sale of bid form</b>	<b>Last date &amp; time of receipt of bid form</b>	<b>Date &amp; time of opening of technical bid</b>
<b>16.11.2021 06.00 p.m</b>	<b>16.11.2021 06.00 p.m</b>	<b>17.11.2021 03.00 p.m</b>	<b>23.11.2021 06.00 p.m</b>	<b>23.11.2021 06.00 p.m</b>	<b>24.11.2021 03.00 p.m</b>

Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

  
**Executive Director (EPM)**  
**RMSCL, Jaipur**