

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur

Phone No.-2223887, Fax no.-0141-2228065

E-Mail : edepmrmisc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: www.rmhc.health.rajasthan.gov.in

F-8()RMSC/EPM/M-4/NIB-654/2021-22/1717

Date:- 23/02/2022

Clarification/Corrigendum/Addendum

Subject:- Amendment in technical specification of item LED O.T. LIGHT DOUBLE DOME No. F-8()RMSC/EPM/M-4/NIB-654/2021-22/1491 Dated: 23.12.2021


In Reference to subject cited above and NIB-654 the various representations received from the firms and issues raised by the Bidders are under consideration by the competent Authorities and technical committee. The following Corrigendum/Addendum is issued for inclusion in bid document as below:-

A) Revised Technical Specifications of LED O.T. LIGHT DOUBLE DOME

1. Surgical light unit with Electronics control , for homogeneous and shadow less light with minimum temperature rise.
2. The highly counter balanced mechanism to fix the dome at any position, easy maneuverable free up & down & swing movements & freely rotatable at any angle(360 degree)
3. Light intensity should be adjustable from control panel available on dome. **Wall mount control panel/Wireless remote control shall be provided.**
4. Focusing should be adjustable from control panel or manually by sterile handle. Sterile handles two numbers extra should be supplied with light .
5. Light intensity at a distance of 1 meter for each dome at least 1,60,000 lux and Light field diameter shall be above 20 cm or better.
6. Life of LEDS MINIMUM 40,000 working hrs or more.
7. Color temperature variable (3800 to 4500 degree k)
8. Light dome shall be compatible for laminar air flow.
9. Power Supply: 230 Volts AC, 50 Hz.
10. Manufacturer should be ISO 13485 certified.(Valid documentation should be submitted in technical bid.)
11. Equipment must have safety requirement of IEC 60601-1-1 and IEC 60601-1-2-41/**European CE certified from notified body/ USFDA(510k) approved. (Documentation should be submitted in technical bid)**
12. Installation will be done free of cost by supplied with all Civil, Mechanical and Electrical works required.
13. Guarantee: Three years from the date of Installation.
14. CMC: CMC will be given @ 5% plus GST (as applicable) and yearly escalation of 5 % on last year's CMC price . The CMC may be awarded for five years (on Yearly basis) after guarantee period of three years, if required.
15. The Service provider should be based in Rajasthan.
16. The Company should Submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the Technical bid.
17. Final technical approval after demonstration.

Handwritten signatures

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.


Executive Director (EPM)
RMSCL, Jaipur