



**RAJASTHAN MEDICAL SERVICES CORPORATION LTD.**

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011ISGC035067

E-Mail - [edepmrmssc-ri@nic.in](mailto:edepmrmssc-ri@nic.in)

No. F-8( )RMSC/EPM/M-7/2020-21/NIB-553/ 267

Dated: 15/12/2020

**CLARIFICATION/CORRIGENDUM/ADDENDUM**

**Subject: Revised NIB Detail, Technical Specification and Amedment in BOQ of item Vaccine Carrier , under No. F-8( ) RMSC/EPM/M-7/2020-21/NIB-553/236 date: 01.12.2020**

In reference to subject cited above NIB-553, various representation received from the firms and issues raised by the bidder are examined by the competent authorities and technical committee. The following clarification/corrigendum/addendum is issued for inclusion in bid document of items as below:-


**A. Revised NIB Detail:-**

| S . N . | Description of Goods under Procurement | Bid Value (in Rs.) | Indicative Quantity (Number) | Minimum average Gross annual turnover for last three financial years (In Rs.) | Minimum average gross turnover for last three financial years for (MSMEs of Rajasthan) (In Rs.) | Bid Security (In Rs.) | Bid Security for MSME Units of Rajasthan (In Rs.) |
|---------|--|--------------------|------------------------------|---|---|-----------------------|---|
| 1.      | Vaccine Carrier with 04 ice packs      | 1,80,00,000        | 30000                        | 10800000  | 5400000   | 180000                | 45000   |
| 2.      | Ice Pack                               | 15,00,000          | 50000                        | 2500000   | 250000  | 15000                 | 3750  |

**B. Revised Technical Specification of Vaccine Carrier -**

1. Vaccine carriers for safe transportation of vaccine from health institution to session site. It should be portable and easy to carry.
2. It should maintain the inside temperature between +2°C to +8°C for 12 hours with four conditioned ice packs if not opened frequently.
3. It should carry minimum 16-20 vials with diluents from health institution to session site.
4. It should have vaccine storage capacity 1.6 to 2.0 liters.
5. It should have weight less than 5 kg when fully loaded.
6. It should have weight 2.5 kg to 3 kg with empty ice packs.
7. It should have HDPE (High Density polyethylene) for external surface and internal lining.
8. It should have insulation of CFC free polyurethane, (PUF) of 40-50 m.m. thickness minimum.
9. It should have hold over time of 30-40 hrs at 43°C ambient temperature.
10. Ice packs of WHO PQS performance specifications. Ice packs should contain water 0.35 to 0.40 ltr. and external dimensions of Ice packs should be 163x94x34 mm ±2 m.m. Ice packs should be of HDPE and it should pass the robust and leakage test. Ice packs should have removable caps for water filling and also have a filling line mark clearly indicated. Ice packs when frozen, the thickness should not exceed 25% of unfrozen Ice packs thickness.

*15/12/2020*



11. A soft foam pad of 30 m.m thickness to be provided inside vaccine carrier as a temporary lid.
12. A rectangle/circular pan (plastic) of suitable size to be provided in between icepacks to carry the vaccine vials so that vaccines vials may not come in direct contact with Ice packs.
13. The equipment should conform to WHO PQS performance specification.
14. The manufacture must provide warranty for two year for replacement.
15. The manufacture should provide a certificate of compliance of WHO test procedures by an independent ISO/IEC 17025 testing laboratory, accredited for type testing of vaccine carriers by WHO/UNICEF/STQC/NABL.
16. Supplier should provide two years guarantee as per specification.
17. **The rate of ICE pack shall be quoted separately in BOQ.**

**Terms and Conditions:**

1. Guarantee: 2 Years on equipment from the date of installation.
2. The company should mention the Make and Model name/ Number of the quoted equipment and submit the technical brochure of the quoted model in the technical bid along with the compliance sheet as per the technical specification.

**C. Revised BOQ:-**

**BF-4**

**Financial bid for Quoted Item**

| S. N. | Name of item with full specification | Approx qty. | Packing unit | Net rate per packing unit (In Rupees) | Rate of SGST | Rate of CGST | Rate of IGST | Rate of GST as applicable | Amount of GST as applicable | Total amount In Rupees. (6+11) |
|-------|--------------------------------------|-------------|--------------|---------------------------------------|--------------|--------------|--------------|---------------------------|-----------------------------|--------------------------------|
| 1     | 2                                    | 3           | 4            | 5                                     | 6            | 7            | 8            | 9                         | 10                          | 11                             |
| 1     | Vaccine Carrier with 04 ice packs    | 30000       | Each         | Do not quote rates here.              |              |              |              |                           |                             |                                |
| 2     | Ice packs                            | 50000       | Each         |                                       |              |              |              |                           |                             |                                |

Rates shall only be filled in BOQ <https://eproc.rajasthan.gov.in>

Date

Signature

Name in capitals  
Company /Firm seal

**Note: -**

1. The Bidder should quote the rates in the BOQ provided online, any deviation in quoting rates may lead to rejection of the financial bid.
2. Rates of SGST, CGST & IGST in %, should be shown separately in column no. 6, 7 & 8 for further reference.
3. Rate should be quoted only for packing units as mentioned in the bid.
4. No quantity or cash discounts should be offered.
5. Read all the terms & conditions before filling the BOQ online.
6. Please quote rates in absolute amount only.
7. The L-1 will be decided on the basis of Base rate per unit quoted in online BOQ for each goods.
8. The rates of consumables, accessories etc. quoted by the bidder separately shall be considered only for the purpose of deciding lowest rates of consumables, accessories etc. of the successful bidder.
9. Specification of Ice packs should be same as per Point no 10 of Tech. Specifications of Vaccine Carrier.




9. Specification of Ice packs should be same as per Point no 10 of Tech. Specifications of Vaccine Carrier.

It is also clarified that information of award of contract shall be communicated to all participating bidders on the website [www.rmssc.nic.in](http://www.rmssc.nic.in) and [sppp.raj.nic.in](http://sppp.raj.nic.in). Please note that individual bidder will not be intimated.”

**Note:-** Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

This bears the approval of M.D., RMSCL, Jaipur.

  
**Executive Director (EPM)**  
**RMSCL, Jaipur**

