



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.
D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail - edepmrm-sc-rj@nic.in

No. F-8()RMSC/EPM/M-7/2020-21/NIB-540/ 313

Dated: 08/01/2021

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Revised Technical Specification and revised bid schedule of item " X-ray Machine (Dental)" No. F-8() RMSC/EPM/M-7/NIB-540/2020-21/139 date: 21.10.2020.

In reference to subject cited above NIB-540, various representation received from the firms and issues raised by the bidder are examined by the competent authorities and technical committee. The following clarification/corrigendum/addendum is issued for inclusion in bid document of items as below:-

Revised Technical Specifications of Radiology X-ray Machine (Dental)

DENTAL X-RAY UNIT **60-70 KV**, upto 7 mA or more (A fine step change in KV and mA is required)

60-70 KV, up to 7 mA or more (AERB approved), with digital timer

High frequency DC range 200 khz and above

Dental X-Ray built in automatic arch timer and timer can be adjusted manually, if required Mode for adult, Children & RVG

Specification:-

- 1) Operating voltage: 220-230 VAC - 50HZ (+/-10%)
- 2) Exposing time: 2 Sec.(Max)/0.05 or less Sec(Min.)

Features:-

- Digital display provided for Exposure time setting
- Easy exposure time set control by soft touch buttons.
- Automatic power cut off at low/high voltage as a safety measure for X-Ray tube head.
- Microprocessor based controller for setting accurate exposure time.
- X-Ray tube angulation.
- X-Ray tube can be moved to any position convenient to the patient with the help of Arm and Hinge arrangements. The X-Ray tube can also rotated around its axis both vertically (270 degree) and horizontally (360 degree) to help selection of correct angulation.

Technical Data:-

Frequency :50/60Hz

Input current: 220/240V

Tube Voltage:**60-70 KV**

Anodic current: up to 7 mA or more

Total filtration: 2 mm A1 or more

Focal film dis:20 cms.

Focal spot:0.4 to 0.8 mm

X-Ray should have USFDA approval or should have CE Mark issued by a notified body registered in European Commission or **BIS Certificate, under relevant category**. Related documents shall be submitted along with technical bid.

AERB certificate for **offered/quoted** model is must. Related documents shall be submitted along with technical bid.

- a) The unit should be provided with DAYLIGHT DEVELOPING SYSTEM:

Signature

- b) Manual Day light X-Ray developer: It is a compact developing system with viewing window.
c) The total body construction is of non-corrosive fiber glass or other superior material.
Supplied with:-
i) Non corrosive tanks for solutions - 3 Nos.
ii) Hanger of 10 clips - 1 No.
iii) Loose clips - 5 Nos.
iv) There should be guarantee of Three Years.
d) CMC shall be given @ 4 % for 5 years after guarantee period with 5 % escalation every year.
e) Demonstration of the equipment should be required.
f) The Firm shall provide guideline and support to consignee, required to register online official Website eLORA(e-Licensing of Radiation Application).


Revised bid schedule:-

E-bids are invited as per following revised time schedule:-

Existing Dates			Extended Dates		
Last date & time for sale of bid form	Last date & time of receipt of bid form	Date & time of opening of technical bid	Last date & time for sale of bid form	Last date & time of receipt of bid form	Date & time of opening of technical bid
11.01.2021 11:00 a.m.	11.01.2021 6:00 p.m.	12.01.2021 03:00 p.m.	21.01.2021 11:00 a.m.	21.01.2021 6:00 p.m.	22.01.2021 03:00 p.m.

It is also clarified that information of award of contract shall be communicated to all participating bidders on the website www.rmssc.nic.in and sppp.raj.nic.in. Please note that individual bidder will not be intimated.”

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.


Executive Director (EPM)
RMSCL, Jaipur