



**Rajasthan Medical Services Corporation Limited, Jaipur**

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur

Phone No.-2223887, Fax no.-0141-2228065

E-Mail: [edepmrsc-rj@nic.in](mailto:edepmrsc-rj@nic.in)

CIN:U24232RJ2011SGC035067

Website: [www.rmssc.health.rajasthan.gov.in](http://www.rmssc.health.rajasthan.gov.in)

F-80) RMSC/EPM/M-5/NIB-672/ 334

Date:- 25/07/22

**Clarification/Corrigendum/Addendum**

**Subject:-Revised Technical specification and bid security Details of the rate contract for (i) Slit Lamp (ii) Streak Retinoscope (iii) Schiottz Tonometer UNIT for under NIB No. F-80) RMSC/EPM/M-5/NIB-672/2022-23/212 Date 21.06.2022**

In Reference to subject cited above and NIB-672, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Corrigendum/Addendum is issued for inclusion in bid document as below:-

**Revised Technical specification of Streak Retinoscope**

S.N.	Present tender specification	Suggested Amendments in specifications
1.	5. Should have halogen/LED light source	5. Should have LED light source. LED life should be 50,000 hours or more.
2.	11 . Product should be CE certified.	11.Product should be USFDA/European CE certified from notified body. (Valid documentation should be submitted in technical bid)

**Revised Technical specification of Schiottz Tonometer**

S.N.	Tender Specification	Suggested Specification
1.	2. Product should be CE certified.	2. Product should be USFDA/European CE certified from notified body. (Valid documentation should be submitted in technical bid)

**Revised Bid security Details:-**

S. no.	Item Name	Present Bid value and Bid security			Amended Bid value and Bid Security		
		Bid Value	Bid Security	Bid Security MSME	Bid Value	Bid Security	Bid Security MSME
1	Streak Retinoscope	1500000	30000	7500	2400000	48000	12000
2	Schiottz Tonometer	135000	2700	675	150000	3000	750
3	Slit Lamp	4000000	80000	20000	4000000	80000	20000

*(Handwritten signature)*



Please note that all clarification/amendment/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms & conditions remains the same.

A handwritten signature in blue ink is positioned above the printed name of the Executive Director.

**Executive Director (EPM)  
RMSCL, Jaipur**