



**RAJASTHAN MEDICAL SERVICES CORPORATION LTD.**  
D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)

Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail - edepmrmisc-ri@nic.in

No. F-80) RMSC/EPM/M-6/NIB-688/2022-23/ 370

Dated: 10/8/2022

**CLARIFICATION/CORRIGENDUM/AMENDMENT**

**Subject:** Revised Technical Specification & Revised bid schedule of item  
"Ventilator Basic and Ventilator Advance No. F-8(306) RMSC/EPM/M-6 /NIB-  
688/2022-23/ 287 date: 06-07-2022

**Revised bid Technical Specification:-**

- E-bids are invited as per following revised time Specification:-

**Basic Ventilator –**

TECHNICAL SPECIFICATIONS	AMNDED TECHNICAL SPECIFICATIONS
2. It should have user friendly pop-up menu with rotary encoder knob or both.	Point No 2. Read as:- "It should have user friendly popup menu with rotatory encoder knob and touch screen."
17. Reusable Masks (Large & Medium) with each machine - 01 set each	Point No. 17 Read as "Reusable masks (1 Large & 1 Medium Size) & HME filter 20 with each ventilator."
18.HME fiter 20.	Point no. 18 read as "Reusable/Autoclavable Flow sensor and exhalation valve/expiratory cassette-2 nos. each. The expiratory valve should have 3 years complete replacement (Free of cost)"

**Advance Ventilator –**

TECHNICAL SPECIFICATIONS	AMNDED TECHNICAL SPECIFICATIONS
3. Should have integrated Colored Touch screen 12 inch or more..	Point no 3 read as "Should have integrated colored touch screen of 12 inches or more, with rotatory knob "
4. (a) Should have End Tidal CO2 detection by Volumetric Capnography	Point No 4 (a) read as Should have Inbuilt End tidal CO2 detection by Volumetric Capnography on Main Screen.
7. a) Tidal Volume: 2 ml to 2500 ml	Point No 7 (a) read as "Tidal Volume: "2 ml to 2000 ml or More"
7. f) I: E Ratio: 5:1 to 1:5	Point No 7 (f) read as "I:E Ratio : 4:1 to 1:4"
17. Reusable and autoclavable Flow sensor and	Point no. 17 read as

*af*

*Pravin*



exhalation valve / expiratory cassette - 2 nos. each. The expiratory flow sensor and valve should have 5 years complete replacement (free of cost)	"Reusable/Autoclavable Flow sensor and exhalation valve/ expiratory cassette - 2 nos. each. The expiratory valve should have 3 years complete replacement (Free of cost)
• For Basic Ventilator & Advance Ventilator the Bidder Should have Submit Original Catalogue & Datasheet along with technical Bid.	

(A) Amendment in class 21 of instruction to bidder :-

Written representation from the prospective bidders regarding suggestions/clarifications in respect of the provisions of the bidding documents shall only be accepted up to next working day after the date of pre bid meeting, thereafter representations may not be accepted.

**Revised bid schedule:-**

- E-bids are invited as per following revised time schedule:-

Existing Dates			Extended Dates		
Last date & time for sale of bid form	Last date & time of receipt of bid form	Date & time of opening of technical bid	Last date & time for sale of bid form	Last date & time of receipt of bid form	Date & time of opening of technical bid
16.08.2022 06:00 p.m.	16.08.2022 06:00 p.m	17.08.2022 03:00 p.m	25.08.2022 06:00 p.m.	25.08.2022 06:00 p.m	26.08.2022 03:00 p.m

It is also clarified that information of award of contract shall be communicated to all participating bidders on the website [www.rmssc.nic.in](http://www.rmssc.nic.in) and [sppp.raj.nic.in](http://sppp.raj.nic.in). Please note that individual bidder will not be intimated."

**Note:-** Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

This bears the approval of M.D., RMSCL, Jaipur.

Executive Director (EPM)  
RMSCL  
Jaipur