



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)

Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail - edepmrmisc-ri@nic.in

No. F-80 RMSC/EPM/M-6/NIB-550/2020-21/397

Dated: 12/02/2021



CLARIFICATION/CORRIGENDUM/ADDENDUM

**Subject: Amended Specification, Bid Condition of item 1.Binocular Microscope
2.Monocular Microscope under No. F-80 RMSC/EPM/M-6/NIB-550/2020-21/219 date:
20.11.2020**

In reference to subject cited above NIB-550, various representation received from the firms and issues raised by the bidder are examined by the competent authorities and technical committee. The following clarification/corrigendum/addendum is issued for inclusion in bid document of items as below:-

**A-Revised Technical Specifications of 1.Binocular Microscope
2.Monocular Microscope**

1. Revised Technical Specification of Binocular Microscope

1. It should be seidentopf binocular head microscope with infinity optical system.
2. Head should be inclined at 30 degree and 360 degree rotatable without adjusting screws with interpupillary distance 50-75 mm.
3. Eye piece – It should be paired, acromatic wide field, 10x power.Field of view (FOV) 22mm with diopter adjustment on both the eye pieces.
4. Objective – It should have 4 objectives 4x,10x, 40x and 100x, having numerical aperture 0.1,0.25, 0.60-0.65 and 1.25-1.65 respectively oil immersion objective (100x) should be spring loaded. All objectives should be plan achromatic, antifungal coated infinity and color corrected.
5. It should have reverse quadruple revolving nose pieces to accommodate from objectives with click stops. It should be provided with ribbed grip for easy rotation. It should be mounted on ball bearing mechanism or better system.
6. Stage should be uniformly horizontal mechanical stage 140 mm x 140mm with fine vernier graduations. It should be provided with spring loaded slide holder. It should have substage, vertical co-axial adjustment for slide manipulation. It should have ball bearing arrangement on single handle for both transverse (80mm ± 5mm) and front to back (50mm ± 5mm) movements.
7. Condenser should be Abbe type, Numerical aperture (NA) 1.25 focusable with rack and pinion arrangement incorporating an spherical lens and iris diaphragm. It should have filter holder swing in/out blue filter.
8. System should have built in illumination 3W LED with intensity control with inbuilt protective safety device which with sand fluctuations of voltage from 140 V to 280 V.
9. LED illumination 3W with intensity control knob >10,000hrs bulb lifespan with battery backup of 1 hrs and charging indication.
10. Focusing knob should have coaxial coarse and fine focusing knobs. Fine focusing movements should have sensitivity of 2 micron or less, coarse focus with torque adjustment stop safety arrangement should be there.

[Handwritten signature]
12/02/21



11. System should work on universal power supply 140 V – 280 Volt
12. Wooden box with lock for storage should be provided free of cost.
13. Other essential accessories like oil, cleaning solution, dust cover should be provided free of cost with equipment, Firm should quote rate LED bulbs and objective lenses in financial bid.
14. **Quoted Model should be “CE from Notified Body”/BIS (with update amendment) certified, Related certificate must be submitted in technical bid.**
15. **Standards and Safety:-** Bidder should have ISO 13485 certification for quality standards, and equipments also conforms to the standards for electrical safety IEC 60601- General requirements (related documents must be submitted in the technical Bid.)
16. Warantee: Three years on Equipment from the date of installation.
17. CMC: CMC will be given @ 5 % of Net Rate exclusive of GST (As applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years.
18. Installation will be done by supplier free of cost.
19. The service engineer should be based in Rajasthan.
20. Demonstration of Equipment is must for technical approval.
21. The company should mention the make & model name of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet of the manufacturer as per amended technical specifications Firm should mention all the pre-installation requirements in technical bid

2. Revised Technical Specification of Monocular Microscope

1. It should be monocular compound microscope with LED illumination.
2. It should have 4x, 10x, 40x and 100x, Plan acromatic objectives with anti fungus, 40x and 100x objectives should be spring loaded.
3. It should have Wide field 10x/18mm eye pieces.
4. It should have Single mould sturdy stand with anti-rust materials. Extended base for better stability, Comfort pad covers.
5. It should have Co-axial coarse and fine focusing on ball bearing system for smooth operation. Fine drive movement 0.2mm/rotation.
6. It should have Quadruple revolving nosepiece, ball bearing type with rubber grip.
7. It should have Sub stage rack and pinion mounted abbe condenser NA 1.25 with aspheric lens and should have Iris diaphragm with snap-in frosted blue filter.
8. It should have Monocular viewing body, 45° inclined, 360° rotatable with smooth movement.
9. It should have 200 - 240 V AC voltage supply.
10. It should have 370-390 mm height and it should have built in illumination 3W LED white light with intensity control.
11. **Quoted Model should be “CE from Notified Body”/BIS (with update amendment) certified, Related certificate must be submitted in technical bid.**
12. It should have 1 hours battery back-up.
13. It should have rectangular stage size 135 x 124mm, X/Y travel range 76mm x 50mm. Low drive movement controls, single specimen holder.
14. It should provide one spare LED light source, one cedar wood oil bottle of atleast 100 ml, tissue lens cleaner with at least 10 sheets during installation.
15. Wooden box with lock for storage should be provided free of cost.

०२/२१



<p>16. Standards and Safety:- Bidder should have ISO 13485 certification for quality standards, and equipments also conforms to the standards for electrical safety IEC 60601- General requirements (related documents must be submitted in the technical Bid.)</p> <p>17. Warrantee: Three years on Equipment from the date of installation.</p> <p>18. CMC: CMC will be given @ 5 % of Net Rate exclusive of GST (As applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Warrantee period of three years.</p> <p>19. Installation will be done by supplier free of cost.</p> <p>20. The service engineer should be based in Rajasthan.</p> <p>21. Demonstration of Equipment is must for technical approval.</p> <p>22. The company should mention the make & model name of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet of the manufacturer as per amended technical specifications Firm should mention all the pre-installation requirements in technical bid</p>
<p>(a) Bidders are requested to send with bid, printed descriptive literature (catalogue) of the quoted goods.</p> <p>(b) If bidder had supplied or had a rate contract of quoted goods with any other Govt. institution within one year, he may be asked to provide copies of purchase orders, invoices and rate contracts.</p>

B-Revised Bid Condition:-

S.N.	Existing bid condition/Clause/Term	Revised Bid Condition
1	---	1.वित्त विभाग (G&T) की अधिसूचना दिनांक 01.01.2021 एवं वित्त विभाग (G&T) राज. सरकार के आदेश F.2(1)FD/G&T-SPFC/2017 dated 15.01.2021 के अनुसार जारी नियम/संशोधन इस निविदा पर प्रभावी होंगे।
2.	---	2.वित्त विभाग (G&T) की अधिसूचना दिनांक 18.12.2020 एवं वित्त विभाग (G&T) राज. सरकार के आदेश F.2(1)FD/G&T-SPFC/2017 dated 23.12.2020 के अनुसार जारी नियम/संशोधन अनुसार बिड़ सिक्यूरिटी राशि प्राप्त नहीं करने एवं इसके स्थान पर बिड़ सिक्यूरिटी के संबंध में घोषणा पत्र (Declarations) प्राप्त करने का प्रावधान हेतु संशोधन इस निविदा पर प्रभावी होंगे। Declaration की प्रति सलंग्न है। उक्त Declaration Rs.50/- के Non-Judicial Stamp Paper पर Notarized कराना है।
3	Page no.3,5,6,11,96... Manufacturer/Direct Importer/Authorized Dealers/Authorized Distributors.	Bids are invited from Manufacturer/Direct Importer only.


14/11/24



It is also clarified that information of award of contract shall be communicated to all participating bidders on the website www.rmssc.nic.in and sppp.raj.nic.in. Please note that individual bidder will not be intimated.”

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

This bears the approval of M.D., RMSCL, Jaipur.


Executive Director (EPM)
RMSCL, Jaipur

Form of Bid-Securing Declaration

Date :
Bid No. :
Alternative No. :

To :

We, the undersigned, declare that:

We understand that, according to your conditions, bids must be supported by a Bid-Securing Declaration. We accept that we are required to pay the bid security amount specified in the Term and Condition of Bid, in the following cases, namely :-

- (a) when we withdraw or modify our bid after opening of bids;
- (b) when we do not execute the agreement, if any, after placement of supply/work order within the specified period;
- (c) when we fail to commence the supply of the goods or service or execute work as per supply/work order within the time specified;
- (d) when we do not deposit the performance security within specified period after the supply/work order is placed; and
- (e) if we breach any provision of code of integrity prescribed for bidding specified in the Act and Chapter VI of these rules.

In addition to above, the State Government shall debar us from participating in any procurement process undertaken for a period not exceeding three years in case where the entire bid security or any part thereof is required to be forfeited by procuring entity.

We understand this Bid Securing Declaration shall expire if :-

- (i) we are not the successful Bidder;
- (ii) the execution of agreement for procurement and performance security is furnished by us in case we are successful bidder;
- (iii) thirty days after the expiration of our Bid.
- (iv) the cancellation of the procurement process; or
- (v) the withdrawal of bid prior to the deadline for presenting bids, unless the bidding documents stipulate that no such withdrawal is permitted.

Signed :-----

Name :-----

In the capacity of :-----

Duly authorized to sign the bid for and on behalf of :

Dated on day of

Corporate Seal -----

[Note: In case of a Joint Venture, the Bid Securing Declaration must be signed in name of all partners of the Joint Venture that is submitting the bid.]

