

RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)

Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail - edepmrm-sc-ri@nic.in

No. F-8(194) RMSC/EPM/M-2/2020-21/NIB-459/ 4709

Dated: 4.9.20

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Amended Technical Specification and revised bid schedule of bid document of item Glucostrips, under No. F-8(194)RMSC/EPM/M-2/NIB-459/2020-21/4325 date: 20.06.2020

In reference to subject cited above NIB-459, various representation received from the firms and issues raised by the bidder are examined by the competent authorities and technical committee. The following clarification/corrigendum/addendum is issued for inclusion in bid document of items as below:-

1. REVISED TECHNICAL SPECIFICATION OF GLUCOSE STRIPS

1. Should be based on Glucose oxidase enzymatic method.
2. Glucostrip should not use more than 2 micro liter blood sample for glucose testing.
3. Should be able to use fresh capillary whole blood.
4. All strips should have at least one year expiry from the date of supply.
5. Strips should be packed individually/in a air tight bottle pack.
6. Strips should be available in the local market.
7. Auto disable lancet should be provided free of cost with each strip.
8. Strips should have packing of 100Nos.

Note-

9. Compatible Glucometer should be provided free of cost on purchase of every 1000 strips.
10. Cost of one single strip should be provided in BOQ.

Specification of Glucometer to be provided free of cost with glucose strips-

1. It should be a hand held meter.
2. It should be easy to use, one/two/three button navigation system. It should have LCD display with symbols.
3. It should be based on Electrochemical (Gold Electrode)/Biosensor principal.
4. It should give fast and accurate result within 5-8 seconds (blood application with test strip within the meter)
5. No Coding system (Automatic calibration).
6. It should have under dosing detection system when less blood applied on the test strip.
7. Sample size should be 1-2 uL.
8. It should have measuring range between 20-500 or more mg/dL of blood sugar.
9. It should work in the temperature range +5°C to 45 °C and up to 85% humidity.
10. It should have Lithium coin battery.
11. It should have up to 200 test memory with date & time.
12. Final approval after technical demonstration.
13. Guarantee shall be 3 year on Glucometer and 06 months on Battery.
14. License of manufacturing from Drug Authority/Import License from Drug Authority, Renewal/Retention copy and Product approval/Permission copy, related documents should be submitted in Technical Bid.
15. It should have valid USFDA/European CE(notified body) with IVD approval/IS-15197:2013 certificate, related documents should be submitted in Technical Bid.
16. BIS Standards are freshly implemented (Order No-HCT/13-14/Medical Device mgmt Model/10 Dated-24th July 2018).
17. At any time if results are not found satisfactory of supplied items the rate contract may be cancelled.
18. Sample should be submitted along with Technical bid for Demonstration.

Other terms and condition to be covered during guarantee period by the supplier:-

- Response time - < 48 Hours after logged the complaint in complain portal.
- Service hours - As per health facility schedule.
- Preventive Maintenance Schedule (PM)-As per OEM (Mention PM Schedule)
- Calibration Schedule – As per OEM (Mention Calibration Schedule)
- Up time - 95% (346 Days)
- Technical & Application Support, Demonstrations & Trainings - As & when required
- Toll Free No-
- Life of the equipment- As per OEM

2. **Revised Bid Schdeule--**

E-bids are invited as per following revised time schedule:-

Existing Dates		Extended Dates	
Last date & time of receipt of bid form	Date & time of opening of technical bid	Last date & time of receipt of bid form	Date & time of opening of technical bid
<u>10.09.2020</u> <u>6:00 p.m.</u>	<u>11.09.2020</u> <u>11:00 a.m.</u>	<u>24.09.2020</u> <u>6:00 p.m.</u>	<u>25.09.2020</u> <u>11:00 a.m.</u>

Please note that all clarification/amendment/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms & conditions remains the same.


Executive Director (EPM)
RMSCL, Jaipur