



Rajasthan Medical Services Corporation Limited, Jaipur
Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur

Phone No.-2223887, Fax no.-0141-2228065

E-Mail: edepmrmisc-ri@nic.in

CIN:U24232RJ2011SGC035067

Website: www.rmsc.health.rajasthan.gov.in

F-8(RMSC/EPM/M-4/NIB-494/2020-21/4783

Date:- 14.9.20

Clarification/Corrigendum/Addendum

**Subject:-Revised Technical Specification and Date Extension for NIB No. F-8()
RMSC/EPM/M-3/NIB-494/2020-21/4485 Dated 24.7.2020 for item 1.Digital
Watch 2.Digital Thermometer 3. Neonatal Weighing Scale-Tubular and
4.Sling of the weighing scale.**

In Reference to subject cited above and NIB-494, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Corrigendum/Addendum is issued for inclusion in bid document as below:-

1. **Revised Technical Specification of HBNC Kit**
1. **Digital Watch:**
 - a) Date and time, (in Hours, Minutes and seconds)
 - b) Illuminated dial
 - c) Standard Numerals (not Roman or others) - stop watch function is essential
 - d) Battery life of at least one year
 - e) Readily available battery
2. **Digital Thermometer:**
 - a) Range of temperature measurement 32°C to 42°C (89.6°F to 109.4°F)
 - b) Can be calibrated in both Centigrade and Fahrenheit, but if only one option is available, then Fahrenheit is preferable.
 - c) Easy to read LCD display
 - d) Buzzer signal function
 - e) Takes 60-90 seconds to measure temperature
 - f) Can be used in the armpit/axilla, orally and rectally
 - g) Accuracy of temperature $\pm 1^\circ\text{C}$ and $\pm 0.2^\circ\text{F}$
3. **Neonatal Weighing Scale-Tubular Spring Type with Sling:**
 - a) Hanging tubular handheld type
 - b) Plastic tubular body
 - c) Able to weight between 0 to at least 5000 grams
 - d) Clear and easy to read measurement panel with specific colour coded graduation at 100 grams difference (green- >2500 grams, yellow- 2000-2500 grams and red- <2000 grams)
 - e) Zero adjustment facility
 - f) Corrosion protected load hook and suspension ring/hook: at top for holding while weighing and at lower end for handing the baby holding sling.
 - g) Soft, durable sling material for holding the baby
 - h) Warrantee - 3 years

Handwritten signature

4. **Sling of the Weighing Scale:**


- a) Sling should be made of parachute cloth with minimum 50 GSM (grams per square meter) and relevant document (test report) shall submit along with technical bid. Cloth and stitching of the sling should be smooth and soft for the baby.
- b) Size of the sling should be big enough to hold a newborn and maximum weight bearing capacity of 10 kg. (Approximate length- 74cms and breadth-54cms)
- c) Sling should have two loops made of nylon minimum 1 inch width one at each end and which can be hanged in the hock of spring scale.

2. **Date extension of Bid:-**

Please note that all clarification/amendment/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

Previous Last Date & time of receipt of Bid Form	New Last Date & time of receipt of Bid Form	Previous Date & time of opening of technical bid	New Date & time of opening of technical bid
14.09.2020 06:00 PM	21.09.2020 06:00 PM	15.09.2020 11:00 AM	22.09.2020 11:00 AM

All other terms & conditions remains the same.


Executive Director (EPM)
RMSCL, Jaipur