



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.
D- Block, SwasthyaBhawan, TilakMarg, C-Scheme JAIPUR (Raj)

Ph. No. 0141-2223887, Fax No. 0141-2228065 CIN: U24232RJ2011SGC035067 E-Mail -edepmrmsc-ri@nic.in

No. F-8() RMSC/EPM/M-6/NIB-702/2022-23/492

Dated: 29/9/22

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Revised Technical Specification of item "Weighing Scale Adult (Digital)", under No. F-80RMSC/EPM/M-6/NIB-702/2022-23/ 417 date: 29.08.2022

Revised Technical Specification --

E-bids are invited as per following revised Specification :-

1. Should be Digital type weighing machine.
2. Should have LED/LCD screen to display weight.(With min. height of display is 14 mm.)
3. Range of weighing: 2 to150 kg.or more
4. Should have auto Zero facility.
5. Size of scale- 12x12 inch.
6. Should have accuracy and resolution of 100 gram.
7. Machine should have Auto-OFF.
8. The Scale should have inbuilt battery.
9. The manufacturer shall have the valid license and should have model approval by the legal metrological department and the weighing scale must be stamped by the legal metrological department. In case of Importer, the bidder should have valid Importer and repair license from legal metrological department, Govt. of Rajasthan.
10. Manufacturer must have ISO 9001.
11. Final technical approval after demonstration.
12. Guarantee: ONE year on full product (including battery) from the date of installation.

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/addendum should be signed and annexed with bid document. All other terms & conditions remains the same.


Executive Director (EPM)
RMSCL, Jaipur