



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, SwasthyaBhawan, TilakMarg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065 CTN: U24232RJ2011SGC035067 E-Mail -edepmrmisc-rj@nic.in

No. F-8() RMSC/EPM/M-3/22-23/NIB-700/502

Dated: 7/10/22

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Amendment in technical specification of bid document of item 3 Channel ECG Machine for under NIB No. F-8() RMSC/EPM/M-3/NIB-700/2022-23/405 dated 25.08.2022

In reference to subject cited above NIB-700, various representation received from the firms and issues raised by the bidder are examined by the competent authorities and technical committee. The following clarification/corrigendum/addendum is issued for inclusion in bid document of items as below:-

1. Revised Technical Specification item 3 Channel ECG Machine:-

- 1) 3 Channel ECG machine with colour TFT/LCD display. Size 5" or more.
- 2) Automatic, Manual Recording Modes.
- 3) Soft touch keyboard.
- 4) Paper size: 60 mm or more.
- 5) Should have visual alarm for open lead.
- 6) Mains and Rechargeable battery operated.
- 7) Inbuilt Rechargeable battery for at least 4 hours.
- 8) Portable, lightweight
- 9) Measurement facility with interpretation.
- 10) Memory for 100 ECG or more.
- 11) Automatic Lead Marking. Thermal printing.
- 12) Recording Sensitivity: 2.5, 5, 10, 20 mm/mV.
- 13) Defibrillation Protection: Internal
- 14) Frequency Response of 0.05-150 Hz.
- 15) Filters: Mains and Muscle interference.
- 16) It should have USB Connection.
- 17) It should have pacemaker detection.
- 18) Safety standards Class II type CF
- 19) Quoted model should be European CE (notified body) or USFDA approved.
- 20) Guarantee 3 years on equipment from the date of Installation & minimum 06 months on Consumables.

(Handwritten signature)

- 21) CMC shall be given 5% of net rate (Exclusive GST as applicable) plus service tax applicable and escalation of 5 on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.
- 22) Equipment should be supplied with compatible standard paper rolls (quantity 10 nos), additional Clamp Electrodes (1 set), Chest Electrodes (1 set), ECG Jelly (250 mlX2 bottles), ECG Cables (2 set), Power Cable (1 vet) Earth Cable-1, Guarantee Card & Operating Manual.
- 23) The machine should be supplied with a good quality carry bag to carry ECG Machine
- 24) Detail of Service Network of Firm in Rajasthan.
- 25) The Company should mention the model name/number and submit the technical brochure of the quoted model in the Technical Bid.
- 26) Paper Speed 12.5/25/50/100 mm/sec.

2. Revised Bid Schdeule—

E-bids are invited as per following revised time schedule:-

Existing Dates				Extended Dates			
Last date & time for sale of bid form	Last date & time of receipt of bid form	Last date of submission of bid document fees	Date & time of opening of technical bid	Last date & time for sale of bid form	Last date & time of receipt of bid form	Last date of submission of bid document fees	Date & time of opening of technical bid
<u>10.10.2022</u> <u>11:00 a.m.</u>	<u>10.10.2022</u> <u>6:00 p.m.</u>	<u>11.10.2022</u> <u>03:00 p.m.</u>	<u>11.10.2022</u> <u>03:00 p.m.</u>	<u>20.10.2022</u> <u>11:00 a.m.</u>	<u>20.10.2022</u> <u>6:00 p.m.</u>	<u>21.10.2022</u> <u>03:00 p.m.</u>	<u>21.10.2022</u> <u>03:00 p.m.</u>

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

Executive Director (EPM)
RMSCL, Jaipur