



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)

Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail - edenmrmsc-ri@nic.in

No. F-8(328)RMSC/EPM/M-2/2023-24/NIB-768/ 503

Dated: 14/9/23

CLARIFICATION/CORRIGENDUM/ADDENDUM


Subject: Revised Technical Specification of item " Incubator ", under No. RMSC/EPM/M-2/2023-24/NIB-768/240 Date 14.07.2023.

Revised Technical Specification of Item "Incubator "

Existing Technical Specification in Point No. 17	Revised Technical Specification in Point No. 17
Model test report of IEC 60601-1: Safety requirement for electrical equipment for measurement, control and laboratory use; - from NABL accredited laboratory under scope.	Model test report of IEC 61010-1/ IEC 60601-1: Safety requirement for electrical equipment for measurement, control and laboratory use; - from NABL accredited laboratory under scope.

Please note that all clarification/amendment/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms, conditions & Specification remains the same.


Executive Director (EPM)
RMSCL, Jaipur
