



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.
D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail - edepmrmsc-ri@nic.in

No. F-8()RMSC/EPM/SM-3/2020-21/NIB-523/ 504

Dated: 19.3.21

CLARIFICATION/CORRIGENDUM/ADDENDUM

Following amendment to be incorporated in the BID Document of NIB No. 523/2020-21 for procurement of Dialysis Machine:-

1- Amended Technical specification.

Sr.No.	Existing Technical Specification	Amended Technical Specification
1	It should have in-built battery backup of minimum 30 minutes for complete extracorporeal blood system during power failure.	It should have in-built Lithium Polymer battery backup of minimum 30 minutes for complete extracorporeal blood system during power failure.

Rest of the terms & Conditions shall remain unchanged .


Executive Director (EPM)
RMSCL, Jaipur