

RAJASTHAN MEDICAL SERVICES CORPORATION LTD.
D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail - edepmrmsc-ri@nic.in

No. F-8() RMSC/EPM/M-1/2020-21/NIB-517/ 5256

Dated: 10/11/2020

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Revised Technical Specification of Bid document of item BIPAP Ventilator with well cushioned face mask and harness , under No. F-8() RMSC/EPM/M-1/2020-21/NIB-517/54 date: 26.09.20

In reference to subject cited above NIB-517, various representation received from the firms and issues raised by the bidder are examined by the competent authorities and technical committee. The following clarification/corrigendum/addendum is issued for inclusion in bid document of items as below:-

Revised Technical Specifications BIPAP Ventilator with well cushioned face mask and harness:

1. Non-invasive ventilation support system to ensure proper ventilation to COPD patients.
2. It should be suitable for Adult and Paediatric patients.
3. Modes of operation: CPAP, Spontaneous, S/T, Time Mode & Pressure Control.
4. It should have display to monitoring facility for Tidal Volume, RR, Minute ventilation, I: E ratio.
5. Operating pressure range of IPAP, EPAP & CPAP : 4 to 30 cm H₂O
6. Unit should be capable of delivering Tidal Volume 1200 ml or more.
7. Breath rate should be able to set up to 30 BPM with Ti of 0.3 to 3 sec. and rise time 100ms to 600ms.
8. Sensitivity settings: should have trigger and cycle setting.
9. Unit should have integrated alarm for power, Pressure (High/Low), Low Minute ventilation, Apnea, Respiratory rate (High/Low), Circuit disconnected etc. With digital technology for improved leak compensation during inspiration & exhalation.
10. Complete unit should be supplied in good quality bag which can carry both adult and paediatric circuit and mask along with equipment.
11. Device should have USFDA approval / CE certificate issued by a notified body registered in the European commission, under relevant category. Related documents / certification shall be submitted along with technical bid.
12. Unit should have battery backup capacity of 2 hour or more.
13. Unit should be supplied with reusable patient circuit and reusable full face mask with harness Adult & Paediatric (1 no. each), operational / training manual (1 no.).
14. **Guarantee:** Three years from the date of installation.
15. **CMC:** CMC shall be given @ 5% of net rate (excluding GST as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.
16. Installation will be done by supplier free of cost.
17. Firm should have service engineer network in Rajasthan.
18. The firm should submit technical compliance sheet along with catalogue as per technical specifications mentioning the make & model of quoted item in the Technical bid.

Note:- 1. It is also clarified that information of award of contract shall be communicated to all participating bidders on the website www.rmssc.nic.in and sppp.raj.nic.in. Please note that individual bidder will not be intimated."

2. Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

Executive Director (EPM)
RMSCL, Jaipur