



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.
D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail -edepmrmsc-rj@nic.in

No. F-8()RMSC/EPM/M-6/2020-21/NIB-525/ 5391

Dated: 26/11/2020

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Amended Specification, Bid Condition and Time extension of item Multi Para Monitor of NIB No.F-8()RMSC/EPM/M-6/2020-21/NIB-525/130 dated 19.10.2020

In reference to subject cited above NIB-525, various representation received from the firms and issues raised by the bidder are examined by the competent authorities and technical committee. The following clarification/corrigendum/addendum is issued for inclusion in bid document of items as below:-

A-Revised Technical Specifications of Multi Para Monitor


1. It should be a modular monitor having ECG, SpO2, NIBP, and Respiration, 2IBP and Temperature as standard parameters.
2. Should have facility to display 7 lead ECG, RR, HR, SpO2, NIBP, and Temperature.
3. Monitor should have option for modular upgradable for CO (Cardiac Output), EtCO2 (Capnography).
4. Should display 6 or more waveforms of selected parameters simultaneously, should have colour coding for different waveforms.
5. Should have inbuilt battery backup at least two hour.
6. Display: Color LED touch screen display size of 12" or more.
7. Monitor should have both touch screen and knob to operate.
8. Should have dual temperature monitoring either in 0Celsius or 0Fahrenheit.
9. Should have facility for displaying multi - screen configurations.
10. Should be able to store & display at least 72 hours of graphical trends of all parameters.
11. Should be suitable for monitoring adult & pediatric & neonate patients.
12. The SpO2 technology should sense at low perfusion index in hypotension, shivering motion like Masimo/Nellcor/TruSignal.
13. Should have oscillometric technology for measurement of NIBP with Auto, START and Manual modes.
14. Should have different patient type selection.
15. The pulse rate should be displayed either with ECG or SpO2.
16. The respiration rate should be calculated through Impedance method.
17. Should be able to analyze arrhythmias & ST segment changes.
18. Memory should not wipe off when the power is turned off.
19. Should be able to give visual & audible alarms with three levels of volume adjustment.
20. Should have facility of connecting to the central monitoring station having facility for real time monitoring through Ethernet cable or Wireless Connectivity.
21. Wall mount stand should be provided free of the cost with the monitor.
22. Quoted model should be European CE (Notified Body)/USFDA approved.
23. Warrantee-Three years on equipment from the date of installation.
24. CMC: CMC will be given @ 5 % of Net Rate exclusive of GST (As applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Warrantee period of three years.

(Handwritten signature)

25. Equipment should be supplied with compatible 5 lead ECG cable (1 Nos.), IBP Cable (2 Nos.), SpO2 (Adult) probe with extension cable (2 Nos.), NIBP Cuff with extension cable of Large Adult, Adult & Pediatric (2 each), Temperature probe (2 Nos.) & Operating / user Manual (1 No.)
26. Cost of optional/upgradable parameters should be mentioned separately in BOQ.so that if later stage upgradation is required it can be done.

B-Revised Bid Condition:-

S.N.	Existing bid condition	Revised Bid Condition
1	<p>Page No.67 BF-7(<u>Statement of past supplies and performance</u>)</p> <p>2. Firm should have supplied, installed and commissioned (if required) at least 10% of the indicative quantity of the goods of similar specifications, for which the bid has been submitted in last 36 months i.e. up to the previous month of the date of floating of NIB (the supplies made in the month of floating of NIB shall not be considered).</p> <p>3. The different variants of the goods of the similar specifications (by same manufacturer) may be considered. (E.g.: X company manufactures ECG machines of different types viz., A and B which further come under three variant types viz. aa, bb, and cc respectively. If a bidder firm furnishes documents establishing supply of any type of ECG machine of X company of any type/ variant, it shall be considered as a valid past experience.</p>	<p>BF-7(<u>Statement of past supplies and performance</u>)</p> <p>2. Bidder should have supplied, installed and commissioned (if required) at least 10% of the indicative quantity of the goods under procurement in last 36 months.</p> <p>3.The different variants of the goods under procurement shall be considered. (E.g.: ECG machines of different types viz., A and B which further come under three variant types viz. aa, bb, and cc respectively. If a bidder firm furnishes documents establishing supply of any type of ECG machine of of any type/ variant, it shall be considered as a valid past experience).</p>


26/11/2020

<p>3</p>	<p>SECTION VI : QUALIFICATION AND EVALUATION CRITERIA (QEC) Part-A Page no.100</p> <p>5. Contractual Experience:- The bidder shall be a Manufacturer/Direct Importer who must have manufactured/ imported and supplied/ installed/commissioned goods of similar specifications, in India satisfactorily to the extent of at least 10% of the quantity specified in the NIB in last 36 months. The duration of 36 months will be calculated up to the previous month of submission of the bid. The list of supply/installation/commissioning reports of the goods under procurement by the OEM/ bidder as given in BF-7 should be supported by self attested copies of purchase orders, indents and invoices of goods supplied/installation/commissioning reports.</p> <p>Part-B Page No.101</p> <p>1.Supply experience:- The goods offered/ being procured should have been in production for at least three financial years and a minimum of 10% units should have been sold and have been in operation satisfactorily for at least three years. Spare parts of quoted model shall be available with firm for minimum 10 years (or life span of equipment) for repairing of equipment and shall have to be provided to the RMSCL/ health facility of Rajasthan/ service provider approved by RMSCL/ department/ agency of RMSCL.</p>	<p>SECTION VI : QUALIFICATION AND EVALUATION CRITERIA (QEC) Part-A Page no.100</p> <p>5. Contractual Experience:- As this point is also included in technical criteria it is deleted due to repetition.</p> <p>Part.B Page no. 101</p> <p>1. Supply experience:- The goods offered/ being procured should have been in production for at least three years and Bidder should have supplied, installed and commissioned (if required) at least 10% of the indicative quantity of the goods under procurement in last 36 months.Spare parts of quoted model shall be available with firm for minimum 10 years (or life span of equipment) for repairing of equipment and shall have to be provided to the RMSCL/ health facility of Rajasthan/ service provider approved by RMSCL/ department/ agency of RMSCL.</p>
----------	--	--

20/11/2020

C-Revised bid schedule:-


- E-bids are invited as per following revised time schedule:-

Existing Dates			Extended Dates		
Last date & time for sale of bid form	Last date & time of receipt of bid form	Date & time of opening of technical bid	Last date & time for sale of bid form	Last date & time of receipt of bid form	Date & time of opening of technical bid
01.12.2020 11:00 a.m.	01.12.2020 6:00 p.m.	02.12.2020 03:00 p.m.	10.12.2020 11:00 a.m.	10.12.2020 6:00 p.m.	11.12.2020 03:00 p.m.

It is also clarified that information of award of contract shall be communicated to all participating bidders on the website www.rmssc.nic.in and sppp.raj.nic.in. Please note that individual bidder will not be intimated.”

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

This bears the approval of M.D., RMSCL, Jaipur.


Executive Director (EPM)
RMSCL, Jaipur