



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur

Phone No.-2223887, Fax no.-0141-2228065
CIN:U24232RJ2011SGC035067

E-Mail : edepmrmisc-ri@nic.in
Website : www.rmssc.nic.in

F-8(297) RMSC/EPM/M-3/2017-18/NIB-257/ 5739

Date: 5/12/17

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: -Date extension of Bid document, Revised Technical Specification and other terms and conditions of Bid document for item BIPAP MASK VENTILATOR WITH WELLCUSHIONED FACE MASK WITH HARNESS F-8 (297)RMSC/EPM/M-3/2017-18/NIB-257/5281 dated: 27.10.2017.

In Reference to subject cited above and NIB-257, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Corrigendum/Addendum is issued for inclusion in bid document as below:-

Technical Specifications of BIPAP MASK VENTILATOR WITH WELLCUSHIONED FACE MASK WITH HARNESS

S. No.	Description
1	Technical specifications of BIPAP Ventilator <ol style="list-style-type: none">1. Non-invasive ventilation support system to ensure proper ventilation to COPD patients.2. It should be suitable for Adult and Paediatric patients.3. Modes of operation: CPAP, Spontaneous, S/T, Time Mode & Pressure Control.4. It should have display to monitoring facility for Tidal Volume, RR, Minute ventilation, I: E ratio.5. Operating pressure range of IPAP, EPAP & CPAP : 4 to 30 cm H₂O6. Unit should be capable of delivering Tidal Volume 1200 ml or more when operated in AVAPS or equivalent mode.7. Breath rate should be able to set up to 30 BPM with Ti of 0.3 to 3 sec. and rise time 100ms to 600ms.8. Sensitivity settings: should have trigger and cycle setting.9. Unit should have integrated alarm for power, Pressure (High/Low), Low Minute ventilation, Apnea, Respiratory rate (High/Low), Circuit disconnected etc. With digital technology for improved leak compensation during inspiration & exhalation.10. Complete unit should be supplied in good quality bag which can carry both adult and


- paediatric circuit and mask along with equipment.
11. Unit should be European CE or FDA approved. (Valid certificate should be submitted in technical bid).
 12. Unit should have battery backup capacity of 2 hour or more.
 13. Unit should be supplied with reusable patient circuit and reusable full face mask with harness Adult & Paediatric (1 no. each), operational / training manual (1 no.).
 14. **Guarantee:** Three years from the date of installation.
 15. **CMC:** CMC shall be given @ 5% of net rate (excluding GST as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.
 16. Installation will be done by supplier free of cost.
 17. Firm should have service engineer network in Rajasthan.
 18. The firm should submit technical compliance sheet along with catalogue as per technical specifications mentioning the make & model of quoted item in the Technical bid.

The last date for sale of bid, receipt of bid and date of opening of technical bid hereby extended as below:-

Extended dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
21.12.2017 11:00 AM	21.12.2017 01:00 PM	21.12.2017 03:00 PM

Please note that all clarification/amendment/corrigendum in bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms & conditions remains the same.
This bears the approval of M.D., RMSCL, Jaipur.


5-12-2017
Executive Director (EPM)
RMSCL, Jaipur