



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur

Phone No.-2223887, Fax no.-0141-2228065
CIN:U24232RJ2011SGC035067

E-Mail : edepmrmisc-ri@nic.in
Website: www.rmhc.health.rajasthan.gov.in

No. F-8(313) RMSC/EPM/M-3/2017-18/NIB-273/6006 Date: 26/12/17

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: -Date extension of Bid document, Revised Technical Specifications and other terms and conditions of Bid document for DEEP FREEZER-HORIZONTAL & DEEP FREEZER-VERTICAL of NIB No. F-8 (313) RMSC/EPM/M-3/2017-18/NIB-273/5503 dated: 13.11.2017.

In Reference to subject cited above and NIB-273, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Corrigendum/Addendum is issued for inclusion in bid document as below:-

Technical Specifications of DEEP FREEZER -HORIZONTAL & DEEP FREEZER-VERTICAL

S. No.	Description
1	<p><u>Technical Specifications of Deep Freezer (Horizontal)</u></p> <ol style="list-style-type: none"> Storage capacity 200 litres & Provides temperature of -20°C with accuracy +1°C Outer chamber is made of Mild Steel duly pre treated & finished with powder coated paint for lasting finish. Inner chamber made of STAINLESS STEEL (304 Grade) and should be 100% leak Proof. Outer front double walled door is insulated and is fitted with magnetic tape for air tight closing for no temperature loss, provided with lock and key arrangement. Door operated illumination lamp is fitted inside the chamber for easy visibility. A full view inner plexi-glass door enables inspection and monitoring of inner chamber milk bottles without disturbing the process temperature. High grade PUF INSULATION between outer and inner chamber for minimal thermal losses to maintain -20°C for at least 1 hr in case of power-cut. Direct cooling system for stable temperature control by ISI marked Compressor/Condensing Units CFC free & R-134 a eco friendly refrigerant. Control panel having all controls and circuitry are housed at the top fitted with Temperature controller cum Indicator Microprocessor based with LED display for Set Value (SV) & Process Value (PV) to select any desired temperature. Caster wheel mounted for easy portability. Metal mesh (Jali)for storage of milk containers 30ml ± 5ml. Suitable for operation on 220 to 240 volt, single phase, 50 Hz, A.C. power supply Refrigerant used be non toxic, non hazardous, CFC/HCFC free. Should be provided with Voltage stabilizer or Inbuilt Stabilizer. Should be provided with paper/digital chart recorder. <p><u>Terms & Conditions:</u></p> <ol style="list-style-type: none"> Manufacturer should have ISO 13485 Certificate. Equipment should be ISI/CE certified. Guarantee: Two years on equipment from the date of installation. CMC: CMC will be given @ 5 % plus GST (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for three years (on yearly basis) after completion of Guarantee period of two years. Installation & Training will be done by supplier free of cost.

6. The service engineer should be based in Rajasthan.
7. The company should mention the make & model name / number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.

Technical Specifications of Deep Freezer (Vertical)

1. Storage capacity 200 liters & Provides temperature of -20°C with accuracy +1°C
2. Outer chamber is made of Mild Steel duly pre treated & finished with powder coated paint for lasting finish.
3. Inner chamber made of STAINLESS STEEL (304 Grade) and should be 100% leak Proof.
4. Outer front double walled door is insulated and is fitted with magnetic tape for air tight closing for no temperature loss, provided with lock and key arrangement.
5. Door operated illumination lamp is fitted inside the chamber for easy visibility.
6. A full view inner plexi-glass door enables inspection and monitoring of inner chamber milk bottles without disturbing the process temperature.
7. The door and walls be insulated with sufficient thickness high quality PUF insulation to maintain -20°C for at least 1 hr in case of power-cut.
8. Direct cooling system for stable temperature control by ISI marked Compressor/Condensing Units CFC free & R-134 a eco friendly refrigerant.
9. Control panel having all controls and circuitry are housed at the top/eye level fitted with Temperature controller cum Indicator Microprocessor based with LED display for Set Value (SV) & Process Value (PV) to select any desired temperature.
10. Racks for storage of milk containers 30ml ± 5ml.
11. Caster wheel mounted for easy portability.
12. Suitable for operation on 220 to 240 volt, single phase, 50 Hz, A.C. power supply
13. Refrigerant used be non toxic, non hazardous, CFC/HCFC free.
14. Should be provided with Voltage stabilizer or Inbuilt Stabilizer.
15. Should be provided with chart recorder.

Terms & Conditions:

1. Manufacturer should have ISO 13485 Certificate.
2. Equipment should be ISI/CE certified.
3. **Guarantee:** Two years on equipment from the date of installation.
4. **CMC:** CMC will be given @ 5 % plus GST (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for three years (on yearly basis) after completion of Guarantee period of two years.
5. Installation & Training will be done by supplier free of cost.
6. The service engineer should be based in Rajasthan.
7. The company should mention the make & model name /number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.

The last date for sale of bid, receipt of bid and date of opening of technical bid hereby extended as below:-

Extended dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
10.1.2018 11:00 AM	10.1.2018 01:00 PM	10.1.2018 03:00 PM

Please note that all clarification/amendment/corrigendum in bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms & conditions remains the same.
This bears the approval of M.D., RMSCL, Jaipur.


26.12.2017
Executive Director (EPM)
RMSCL, Jaipur

2

