



Rajasthan Medical Services Corporation Limited

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

CIN:U24232RJ2011SGC035067

E-mail : edepmrmsc-rj@nic.in

Website: www.rmsc.health.rajasthan.gov.in



F-8(216) RMSC/EPM/M-2/NIB-512/2020-21/ 64

Dated: 01/10/2020

Clarification/Corrigendum/Addendum

Subject:-Revised Technical Specification for NIB No. F-8(216) RMSC/EPM/M-2/NIB-512/2020-21/29 dated 18.09.2020 for item Hemoglobin Scale Book.

In reference to subject cited above of NIB-512, various representations received from the firms and issues raised by the bidders are being examined authorities and technical committee. The following corrigendum/addendum is issued for inclusion in bid document as below:-

1. Revised/Amended Technical Specifications :-

1. Simple device for estimating haemoglobin.
2. The Kit should consist of a laminated/water proof booklet containing a set of six shades of red and a 1X200 strips in a weather proof dispenser box that must contain 100 strips on each side of the box having 2 openings dispensing side containing 50X2 (100) strips on each side totalling to 200 strips. A sample must be deposited along the tender.
3. The shades should represent a range of haemoglobin value from 4 to 14 g/dl. By matching the colour of a drop of blood (fresh capillary blood) on a test-strip with one of the shades of red, you can see if the blood is anaemic and, if so, the severity of anaemia in clinical terms.
4. Precise results comparable with the best laboratory methods.
5. It should be CE/ISO (13485-2016) certified. Validation document from Confederation of American Pathologist Accredited (CAP) or NABL accredited must be attached with the tender and the validation must be submitted of the current batch to be supplied by the supplier. The certifying laboratory should have facility/scope of testing/examining the said parameters of the goods under procurement from NABL.
6. The booklet and the dispenser box should be packed in a weather proof packaging box.
7. Shelf life of strips should be minimum three years and date of manufacturing, date of expiry and batch no. must be mentioned on the dispenser box.
7. User manual/ user guidelines in English/Hindi must be provided in the kit/ box.
8. The product can be stored on room temperature. The strip must be kept in their closed Dispenser to avoid becoming damp when not actually being used.
9. Packing of 50 Strips packed with damp proof sealing must be kept in their closed dispenser to avoid becoming damp when not actually being used.

[Handwritten signature]
01/10/2020



2. Following conditions in the special condition of contract (SCC) of bidding document are added:


1. Note -7 mentioned in BOQ is deleted.
2. L-1 bidder will be decided on the basis of lowest rates received for the goods 'Hemoglobin Scale Book' in the online BOQ.

3. Date extension of Bid :-

Previous Last Date & time of receipt of Bid Form	New Last Date & time of receipt of Bid Form	Previous Date & time of opening of technical bid	New Date & time of opening of technical bid
05.10.2020 06:00 PM	12.10.2020 06:00 PM	06.10.2020 03:00 PM	13.10.2020 03:00 PM

Please note that all clarification/amendment/corrigendum in technical specifications/bid conditions/date extension of bid are integral part of the bid document. This corrigendum/addendum should be signed and annexed with bid document.

All other terms & conditions remains the same.


Executive Director (EPM)
RMSCL, Jaipur