



**RAJASTHAN MEDICAL SERVICES CORPORATION LTD.**  
D- Block, SwasthyaBhawan, TilakMarg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065 CIN: U24232RJ2011SGC035067 E-Mail -edepmrmssc-ri@nic.in

No. F-8( ) RMSC/EPM/M-1/NIB-810/2023-24/ 700

Dated: 2/11/23

**CLARIFICATION/CORRIGENDUM/ADDENDUM**

**Subject: Amendment in technical specification and bid schedule of bid document of item (1) Humidifier (2) Oxygen Blender, under No. F-8( ) RMSC/EPM/M-1/NIB-810/2023-24/599 date: 06.10.2023**

**A. Revised Technical Specification of Humidifier-**

- 1) Should be Servo Controlled
- 2) Display of Airway Temperature.
- 3) Should have heater wire adapter, Airway Temperature probe.
- 4) Have temperature settings for invasive and non invasive mode.
- 5) Device should have Audio/visual alarms.
- 6) Humidifier should be European CE (By Notified Body)/ USFDA approved (Valid certificate should be submitted in technical bid).
- 7) Temperature probes (both) should be from same manufacturer or European CE (By Notified Body)/ USFDA approved (Valid certificate should be submitted in technical bid).

**B. Revised Technical Specification of Oxygen Blender:**

Blender should provide precise mixing of air and oxygen with adjustable knob for FiO<sub>2</sub> from 21 to 100%.

- 1) Outlets - Outlet 0-15 lpm (supply pressure at 50 PSI )
- 2) Additional Outlet 0-3 lpm or 0-3.5lpm (supply pressure at 50 PSI)
- 3) Operating temperature range 15 to 25 degree Celsius
- 4) FiO<sub>2</sub> accuracy : 2~ 3% of full scale
- 5) The Oxygen Blender should be European CE (By Notified Body)/US FDA approved (Valid certificate should be submitted in technical bid)

**B. Revised Bid Schdeule—**

E-bids are invited as per following revised time schedule:-

Existing Dates				Extended Dates			
Last date & time for sale of bid form	Last date & time of receipt of bid form	Last date of submission of bid document fees	Date & time of opening of technical bid	Last date & time for sale of bid form	Last date & time of receipt of bid form	Last date of submission of bid document fees	Date & time of opening of technical bid
<u>08.11.2023</u> <u>11:00 a.m.</u>	<u>08.11.2023</u> <u>6:00 p.m.</u>	<u>09.11.2023</u> <u>02:00 p.m.</u>	<u>09.11.2023</u> <u>03:00 p.m.</u>	<u>16.11.2023</u> <u>11:00 a.m.</u>	<u>16.11.2023</u> <u>6:00 p.m.</u>	<u>17.11.2023</u> <u>02:00 p.m.</u>	<u>17.11.2023</u> <u>03:00 p.m.</u>



**Note:-** Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

Executive Director (EPM)  
RMSCL, Jaipur