

RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

E-Mail -edepmrmsc-rj@nic.in

No. F-8(206)RMSC/EPM/SM-3/2019-20/NIB-500/7-8

Dated: 08 10 2020

CIN: U24232RJ2011SGC035067


CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Time extention and Amended and incorporation of formet no. BF-24 under NIB No.F-8(206)RMSC/EPM/SM-3/2019-20/NIB-500/4643 date: 24.08.2020

In reference to subject cited above in NIB-500, some clarification/Corrigendum has been suggested in Pre-bid meeting. After approval in purchase committee is hereby uploading and all the conditions will be as per BID Document No. 500.

A- Revised Technical Specifications of Examination Light (Mobile)

- 1. Mobile light for medical examination.
- 2. Stainless steel stand with 5 anti-static heavy duty industrial grade swivel castors with locking mechanism.
- 3. Articulated arm **should** be **minimum** 105 cm, spring loaded and integrated transformer.
- 4. LED Light: LED life 50,000 hours or more.
- 5. Light intensity at least 50,000 lux or more at a distance of minimum 80 cm
- 6. Natural white light: 3500K 4500 K color temperature
- 7. Power cord length approx 3 meter
- 8. Power requirements: AC 220-240V, 50 Hz
- 9. There should be control panel function:
 - a) On-Off switch
 - b) Intensity adjustable control
- 10. Product should be CE certified. (Valid documentation should be submitted in technical bid)
- 11. Manufacturer should have ISO 13485 certificate for quality standard.
- 12. User manual (Language English/Hindi) guidelines should be provided by suppler along with equipment.
- 13. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the technical bid.
- 14. Service network of firm should be Rajasthan.
- 15.3 years warrantee on complete product from the date of supply against any manufacturing defects.





B-Revised Technical specification of Electric Instrument Sterilizer

- 1. Sterilizer shall have stainless steel seamless shell body.
- 2. Pure 304 grade stainless steel material should be used in entire body of sterilizer such as shell, lid, perforated tray and inner sheets. Inner sheet thickness minimum 0.7mm
- 3. Heating element and plug should be ISI certified.
- 4. Auto cut off feature should be available for thermostat.

5. Category: Large6. Length: 510 mm7. Width: 200 mm8. Depth: 150 mm

Electric Load: Minimum 2 KW
 Power Supply: AC 220V, 50Hz

C-Revised bid schedule:-

• E-bids are invited as per following revised time schedule:-

Existing Dates			Extended Dates		
Last date & time for sale of bid form	Last date & time of receipt of bid form	Date & time of opening of technical bid	Last date & time for sale of bid form	Last date & time of receipt of bid form	Date & time of opening of technical bid
14.10.2020 11:00 a.m.	14.10.2020 6:00 p.m.	15.10.2020 11:00 a.m.	26.10.2020 11:00 a.m.	26.10.2020 6:00 p.m.	27.10.2020 11:00 a.m.

Executive Director (EPM)
RMSCL, Jaipur