



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail - edepmrmsc-ri@nic.in

No. F-8()RMSC/EPM/M-6/2022-23/NIB-718/ 914

Dated: 23/1/23

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Revised Technical Specification and revised bid schedule of item "1. Infant Weighing Scale, 2. Sling of the Weighing Scale, 3. Ari Timer(Digital Watch)", under No. F-8) RMSC/EPM/M-6/NIB-718/2022-23/667 date: 25.11.2022

(I) Revised Technical Specification of "1. Infant Weighing Scale, 2. Sling of the Weighing Scale, 3. Ari Timer(Digital Watch).

Technical specification of Neonatal Weighing Scale (Tubular Spring Type)-

- Hanging tubular handheld type
- Plastic tubular body
- Able to weight between 0 to at least 5000 grams
- Clear and easy to read measurement panel with specific colour coded graduation at 100 grams difference (green->2500 grams, yellow- 2000-2500 grams and red-<<2000 grams)
- Zero adjustment facility
- Corrosion protected load hook and suspension ring/hook: at top for holding While weighing and at lower end for handing the baby holding sling.
- Warranty 1 year
- Manufacturer should have licence from weights and measurement department.**

Technical Specifications of Ari Timer-

- Date and time, (in Hours, Minutes and seconds).
- Illuminated dial.
- Standard Numerals (not Roman or others) - stop watch function is essential.
- Battery life of at least one year.
- Readily available battery.

Technical Specifications of Sling of the Weighing Scale-

- Sling should be made of parachute cloth with minimum 50 GSM (grams per square meter) and relevant document (test report) shall submit along with technical bid. Cloth and stitching of the sling should be smooth and soft for the baby.
- Size of the sling should be big enough to hold a newborn and maximum weight bearing capacity of 10 kg. (Approximate length- 74cms and breadth-54cms)
- Sling should have two loops made of nylon minimum 1 inch width one at each end and which can be hanged in the hock of spring scale.
- Soft, durable sling material for holding the baby.**

(ii) आईएम Neonatal Weighing Scale (Tubular Spring Type) MSME Rajsthan के लिए Reserved है। उक्त आईएम के लिए केवल MSME Unit Rajsthan ही निविदा प्रस्तुत कर सकती है।

[Handwritten signature]




Revised bid schedule:-

E-bids are invited as per following revised time schedule:-

| Existing Dates | | | | Extended Dates | | | |
|--|---|--|---|--|---|--|---|
| Last date & time for sale of bid form | Last date & time of receipt of bid form | Last date of submission of bid document fees | Date & time of opening of technical bid | Last date & time for sale of bid form | Last date & time of receipt of bid form | Last date of submission of bid document fees | Date & time of opening of technical bid |
| <u>24.01.2023</u> <u>11:00 a.m.</u> | <u>24.01.2023</u> <u>6:00 p.m.</u> | <u>25.01.2023</u> <u>02:00 p.m.</u> | <u>25.01.2023</u> <u>03:00 p.m.</u> | <u>02.02.2023</u> <u>11:00 a.m.</u> | <u>02.02.2023</u> <u>6:00 p.m.</u> | <u>03.02.2023</u> <u>02:00 p.m.</u> | <u>03.02.2023</u> <u>03:00 p.m.</u> |

It is also clarified that information of award of contract shall be communicated to all participating bidders on the website www.rmssc.nic.in and sppp.raj.nic.in. Please note that individual bidder will not be intimated.”

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.


Executive Director (EPM)
RMSCL, Jaipur