

Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065 CIN: U24232RJ2011SGC035067

E-mail: edepmrmsc-rj@nic.in Website : www.rmsc.nic.in

No.F-8(52)RMSC/EPM/M-5/17-18/Breast Cancer Device/ 389

Dated : [4] 9] [

UE Lifesciences (India) Pvt Ltd A-102, Universal Paradise, Nanda Patkar Road, Vile Parle (East), Mumbai-400057 info@uelifesciences.com\

विषय :— I—Breast Exam (iBE) System and iBE Scanhead Cartridge की दरें प्रस्तुत करने के कम में। महोदय,

उपरोक्त विषयान्तर्गत लेख है कि राज्य सरकार की स्वीकृति के अनुसार 28 I –Breast Exam (iBE) System का क्रय किया जाना है। जिसके स्पेसिफिकेशन एवं शर्ते निम्नानुसार है। कृपया अपने दरों से सम्बन्धित प्रस्ताव दिनांक 29.09.17 तक प्रस्तुत करें। I –Breast Exam (iBE) System में काम आने वाली iBE Scanhead Cartridge की दरें भी प्रस्तुत करें। I –Breast Exam (iBE) System / iBE Scanhead Cartridge की दर संविदा की जावेगी जिसके आधार पर विभिन्न जिला अस्पतालों में मांग होने पर उपलब्ध करानी होगी।

Specification for i-Breast Exam

- 1. Non-invasive, radiation-free, breast examination.
- 2. Ultra-Portable and light-weight probe. Probe size: 13 x 6.5 x 8.5cm, weight: 0.5kg. Full system with case size: 27 x 26 x 11cm, weight: 2kg.
- 3. Operable with rechargeable battery (1000 to 2000mAh). Recharging Power: Input 110-220V 50/60Hz at 0.3-0.5A and Output 5V at 2A.
- 4. Wireless operation using Bluetooth 2.0 standard protocol.
- 5. Regulatory Clearances & Approvals
 - a. US FDA Cleared Registration# K142926
 - b. CE Marked Certificate# CE-RS-1700-1558
 - c. ISO 13485 Certificate# MDMS 0716 020092
 - d. WHO GMP Certificate# WGMP 0117 020092
- Storage capacity for 100,000 breast examination scans and full reports.
- 7. Custom Windows Application for Breast Examination with capabilities to:
 - a. Perform breast examination.
 - b. Review the breast examination
 - c. Store the breast examination data and
 - d. Print the final report.
- 8. Replay feature enables the physician to replay a breast examination performed in the past, frame by frame.
- 9. Approved for Electromagnetic compatibility (EMC) testy 1409-054EA REV. A Test Standards: IEC 60601—1-2:2007 (3rd Edition) & IEC 60601-1-2:2014 (4th Edition).
- 10. Approved for the following Electrical and Mechanical Safety Tests
 - a. IEC 60601-1 Ed .3 (2005) + Am.1 (2012) = IEC 60601-1Ed.3.1 (2012).
 - b. The following national Deviation were included in the evaluation AAAMI/ANSI ES 60601-1:2005(R) 2012/A1
 - c. EN 60601-1 Ed.3 (2007) + Am.1 (2003).
 - d. CSA C22.2 No. 60601-1: 2014
- 11. Clinical study validation performance characteristics:

- a. Sensitivity is 84%
- b. Specificity is 94%
- c. Negative Predicative Value is 98%
- d. Positive Predictive Value is 60%
- 12. Storage, Transport and Operating conditions:
 - a. Temperature range: 10° C to 50° C
 - b. Maximum relative humidity: 90%
 - c. Atmospheric Pressure range: 700-1060 hPa
- 13. Guarantee: Three years on equipment from the date of installation.
- 14. CMC: CMC shall be given @ 5 % of net rate excluding GST (as applicable). The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.
- 15. The company should mention the make & model name/number of the quoted equipment and submit the technical brochure of the quoted model .

नियम एवं शर्ते :-

- 1. मूल Manufacturing कम्पनी UE Lifesciences Inc., 3711 Market Street, Suite 800, Philadelphia, PA19104 का Authorization Certificate भारत में विक्रय हेतु प्रस्तुत करना होगा।
- 2. Patented Item Certificate.
- 3. Import Export Certificate. (IEC Certificate).

भवदीय

कार्यकारी निदेशक(ई.पी.एम.) राजस्थान चिकित्सा सेवा निगम, जयपुर

No .F-8(52)RMSC/EPM/M-5/17-18/Breast Cancer Device/ **3899** Dated: 14/9/17-18/Breast Cancer Device/ **3899** Dated: 14/9/17-18/Breast Cancer Device/ **3899**

1. ए.जी.एम (आई.टी), आरएमएससी को भेजकर लेख है कि उपरोक्त सूचना SPPP Portal एवं निगम की वेबसाईट पर सूचना अपलोड करवाने हेतु प्रस्तुत है।

> €० कार्यकारी निदेशक(ई.पी.एम.) राजस्थान चिकित्सा सेवा निगम, जयपुर