



## Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur

Phone No.-2223887, Fax no.-0141-2228065

E-Mail : [edepmrmisc-rj@nic.in](mailto:edepmrmisc-rj@nic.in)

CIN:U24232RJ2011SGC035067

Website: [www.rmssc.health.rajasthan.gov.in](http://www.rmssc.health.rajasthan.gov.in)

F-8(RMISC/EPM/M-4/NIB-561/2020-21/ 325

Date:- 12/11/2021

### Clarification/Corrigendum/Addendum

**Subject:- Revised Technical Specification for NIB No. F-8( ) RMSC/EPM/M-4/NIB-561/2020-21/261 Dated 14.12.2020 for item STEM ALCOHOL THERMOMETER.**

In Reference to subject cited above and NIB-561, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Corrigendum/Addendum is issued for inclusion in bid document as below:-

### Revised Technical Specification of Stem Alcohol Thermometer

S. No.	Amended Technical Specification
1.	<ol style="list-style-type: none"><li>1. <b>Power Source:</b> None</li><li>2. <b>Temperature ranges &amp; accuracy:</b> Upper Limit:+50°C Lower Limit:-30°C Accuracy:±01°C</li><li>3. <b>Scale Markings:</b> Easily readable with a minimum space of 1mm between each line: Long lines (with numbers) : for each 10 degrees Short lines: for even number degrees Short lines:for odd number degrees Safe zones for ranges of 0°C to +8°C and -15°C to -25°C to be marked with a green bar.</li><li>4. <b>Maximum relative humidity:</b> 90%</li><li>5. <b>Casing Specification</b> Non-Corrodible, sealed mechanism/Clip type/stand type.</li><li>6. <b>Robustness:</b> To withstand 5 drops from 1 meter and vibration test.</li><li>7. <b>Construction:</b> The glass column must be protected against breakage and strongly supported so that:<ul style="list-style-type: none"><li>• The column cannot be displayed more than 0.5mm vertically with respect to the scale.</li><li>• Alcohol stem column should be firmly fixed to the marking scale.</li><li>• The reading angle is between 80/100° to plane of support plate.</li></ul></li><li>8. <b>Mounting Specifications:</b> Hook to suspend from self or adhesive with 1mm hole to tie with basket.</li><li>9. <b>Application and remark:</b> Used inside refrigerators or freezers in health centers. Can also be packed with vaccines during transport.</li><li>10. Demonstration is must for final approval.</li><li>11. <b>Guarantee:</b> One Year against any manufacturing defect from the date of supply/delivery.</li><li>12. WHO specification reference: E6/TH.3</li></ol>

All other terms & conditions remains the same.

Executive Director (EPM)  
RMSCL, Jaipur