

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph: No. 0141-2223887 Fax: 0141-2228065

E-mail: edepm@rmsc-rj@nic.in

CIN:U24232RJ2011ISGC035067

Website: http://rmsc.health.rajasthan.gov.in

No. F-8() RMSC/EPM/M-2/17-18/NIB-267/5799

Dated: 13/12/17



CLARIFICATION/CORRIGENDUM/ADDENDUM

Sub:- Revised Technical Specifications of item Syringe Infusion Pump and revised bid submission date under NIB No. F-8() RMSC/EPM/M-2/17-18/NIB-267/5488 dated 10.11.17

In Reference to subject cited above and NIB-267, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in Bid document & Technical Specification of items as below:-

A. Amended Technical Specifications of item Syringe Infusion Pump:-


1. Microprocessor controlled syringe infusion Pump.
2. Should have Rate Mode, Time Mode, Dose Mode, and Drug Library for 50 or more drugs.
3. Should Have Facility For Automatic Syringe Size detection for 5 ml, 10 ml, 20 ml, 30 ml, 50ml,
4. Should also have facility to accept any unknown brand of syringe in the form of custom syringe.
5. **ALARMS** : Infusion Completion, Empty, Occlusion, Near Completion, Low Battery & Adjustable Buzzer Volume
6. Adjustable visual and audible alarms.
7. Should have bolus facility with bolus rate from 0.1 to 1200 ml/hr. Anti bolus and auto bolus facility
8. Rechargeable Battery operating time Approx. 6 hours or more at the rate of 5 ml/hr.
9. Should have 90 degree rotation pole clamp convenient for horizontal bar & vertical IV pole.
10. Should have facility for front loading of syringe.
11. Flow rate adjustable from 0.1 ml/hr – 1200 ml/hr (depending on the syringe capacity).
12. Online changing of delivery rate possible.
13. Internal function alarm. Drive disengaged alarm & should have KVO (KEEP VEIN OPEN) Mode. Should have keypad locking facility.
14. Should be CE Certified/USFDA approved.
15. **Guarantee:** Two years on equipment from the date of installation
16. **CMC:** CMC shall be given @ 5% of net rate (excluding GST as applicable) and yearly escalation of 5% on last year's GMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of two years.
17. User manual with trouble shooting guidance should be provided by supplier.
18. **Rates of Accessories and Consumable:** Rates of consumable like Battery & power cord should be provided by bidder in BOQ in e-proc. Portal.
19. Installation will be done by supplier free of cost.
20. Firm should have service engineer network in Rajasthan.
21. Demonstration of equipment is must for technical approval.
22. The firm should submit technical compliance sheet along with catalogue as per technical specifications mentioning the make & model of quoted item in the Technical bid.

B. The bid is re-scheduled as follows:

Existing Dates			Extended Dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
1	2	3	4	5	6
up to 11:00 A.M. 18.12.2017	up to 1.00 P.M. 18.12.2017	from 3.30 P.M. 18.12.2017	up to 11:00 A.M. 28.12.2017	up to 1.00 P.M. 28.12.2017	from 3.30 P.M. 28.12.2017

This corrigendum shall be signed and annexed with bid document. This bears an approval of Managing Director, Rajasthan Medical Services Corporation Limited, Jaipur.


Executive Director (EPM)
RMSCL, Jaipur