



Rajasthan Medical Services Corporation Limited, Jaipur

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail - edepmrmsc-raj@nic.in

No. F-8() RMSC/EPM/M-1/17-18/NIB-317/ 2018/ १०९

Dated: 28/2/18

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: - Amended Technical Specifications of Item Sterilizer of Bid document of item, under No. F-8() RMSC/EPM/M-1/17-18/NIB-317/ 2017/679 date: 14.02.18

In reference to subject cited above and NIB-317, the representations received from the various firms and issues raised by the Bidders, are examined by the competent Authorities and technical committee and approved by purchase committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in Bid document Amended Technical Specification of items as below:-

1. Amended Technical Specifications of Items STERILIZER as below:-

(01) Technical Specification of STERILIZER

- Category – Large
- Length : mm: 510
- Width : mm : 200
- Depth : mm : 150
- Elect Load KW : 2.00
- Body Type : Stainless steel with tray and sterilization forceps: 10 Inches
- ISI/CE Marked

It is also clarified that information of award of contract shall be communicated to all participating bidders on the website www.rmssc.nic.in and sppp.raj.nic.in. Please note that individual bidder will not be intimated."

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

This bears the approval of M.D., RMSCL, Jaipur.

28.2.2018
Executive Director (EPM)

RMSCL, Jaipur