



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.
D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065 CIN: U24232RJ2011SGC035067 E-Mail -edepmrmisc-rj@nic.in
No. F-8() RMSC/EPM/M-1/2020-21/NIB-507/ 140 Dated: 21/10/2020

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Amendment in Technical Specification, BOQ and Notice Inviting Bid for item Blood Gas Analyzer (ABG Machine), under No. F-8() RMSC/EPM/M-1/NIB-507/2020-21/4858 date: 21.09.2020

In reference to subject cited above NIB-507, various representation received from the firms and issues raised by the bidder are examined by the competent authorities and technical committee. The following clarification/corrigendum/addendum is issued for inclusion in bid document of items as below:-

1. Revised Technical Specification of Blood Gas Analyzer (ABG Machine) -

1. Blood Gas Analyzer for the measurement of pH, pO₂, pCO₂, HCT, Lactate, HB and Electrolytes minimum Na, K, Cl and Calcium from whole blood with calculated parameters. All parameters should be in a single unit.
2. Machine should be sensor based/cartridge suitable for multiple tests.
3. Machine should have auto calibration technique.
4. Instruments should have user friendly software.
5. Instruments should have inbuilt memory for patients data & quality control data.
6. Instruments should have battery back-up or online UPS should be supplied with the machine for minimum of half an hour.
7. Analyzer should be able to measure capillary samples as well as syringe samples.
8. Instruments should have inbuilt printer.
9. Cassette/ cartridge should have on board stability of **minimum 28 days or more**.
10. Instruments must be USFDA & CE (notified body) approved, related certificates shall be uploaded in technical bid.
11. System should have comprehensive quality control programme.
12. Company should have own control & calibration.
13. Instruments should have Inbuilt thermal printer & RS 232/ USB port to connect with external computer and the equipment should be supplied with necessary software and hardware.
14. The service engineer should be based in Rajasthan.
15. Equipment should be supplied with one standard consumable pack along with quality control.
16. The company should mention the make & model name/number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.
17. 02 Cassette/Cartridge of 100 test shall be provided free of cost.





2. Revised BOQ:-

Financial bid for Quoted Goods

S. N.	Name of goods with full specification	Brand	Approx qty.	Unit	Net rate per Unit (Rs.)	Rate of SGST (%)	Rate of CGST (%)	Rate of IGST (%)	Rate of GST as applicable	Amount of GST as applicable	Total amount In Rs. (6+11)	
1	2	3	4	5	6	7	8	9	10	11	12	
1.	Blood Gas Analyzer (ABG Machine)		40	Each	Do not quote rates here.							
2.	Cassette/ Cartridge (performing 100 test)		-	100								
Rates shall only be filled in BOQ https://eproc.rajasthan.gov.in												

Signature

Date

Name in capitals

Company /Firm seal

Note: -

1. The Bidder should quote the rates in the BOQ provided online, any deviation in quoting rates may lead to rejection of the financial bid.
2. Rates of SGST, CGST & IGST in %, should be shown separately in column no. 7, 8 & 9 for further reference.
3. Rate should be quoted only for packing units as mentioned in the bid.
4. No quantity or cash discounts should be offered.
5. Read all the terms & conditions before filling the BOQ online.
6. Please quote rates in absolute amount only.
7. The L-1 will be decided on the basis of sum of Base rate per unit quoted in online BOQ of the cost of ABG Machine and cost of 100 Cassette/ Cartridge performing 100 Test [L1=Base Rate of 01qty of ABG Machine + Cost of 100 qty of Cassette/Cartridge (performing 100 Test)]
8. The rates of consumables, accessories etc. quoted by the bidder separately shall be considered only for the purpose of deciding lowest rates of consumables, accessories etc. of the successful bidder.

3. Revised Notice Inviting Bid (NIB-507/2020-21)

Sealed Single Stage Two- envelopes unconditional online Bids for Two Year Rate Contract, are invited on behalf of the Governor of Rajasthan for the procurement of Goods as listed below, from manufacturers/ importers only as per the Bidding Schedule given below:

S.No.	Detail	Information
1	NIB No.	507/2020-21
2	Name of goods	Blood Gas Analyzer (ABG Machine)
3	Estimated Bid value	Rs. 1,40,00,000/-
4	Bid download start date and time	21.09.2020 ; 01:00 p.m.
5	Last date, time of receipt of bid	11.11.2020 ; 06:00 p.m.
6	UBN	


Details of the bidding documents can be accessed or downloaded from the website "sppp.raj.nic.in" or "www.dipronline.org" or "https://eproc.rajasthan.gov.in" or "www.rmhc.health.rajasthan.gov.in".



It is also clarified that information of award of contract shall be communicated to all participating bidders on the website www.rmssc.nic.in and sppp.raj.nic.in. Please note that individual bidder will not be intimated."

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

This bears the approval of M.D., RMSCL, Jaipur.


Executive Director (EPM)
RMSCL, Jaipur

