



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN:U24232RJ2011SGC035067

E-Mail - edepmrmsc-ri@nic.in

No. F-8(205) RMSC/EPM/M-3/16-17/NIB-207/ 451

Date: 2/2/17

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Revised Technical Specification, Minimum average gross annual turnover for last three financial years & required Bid security and revised bid submission date of item Table Post-Martum (Mortuaries) under NIB No. F-8(205) RMSC/EPM/M-3/16-17/NIB-207/6127 Dated 14.12.2016.

In Reference to subject cited above and NIB-207, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Corrigendum/Addendum is issued for inclusion in bid document & Technical Specification of items (Section-V, Schedule of Supply, Point no. 3) as below:-

1. Revised Technical Specification of POST MORTUM TABLE

1. SIZE: Width :- 102" , Depth:- 30" , Height :- 34"
2. TOP: 1.5mm thick S.S. secured to framework. All ends shall be closed, welded, ground and polished smooth. Top should be sound deadened.
3. FRAMEWORK: 38 mm x 38 mm x 3 mm S.S. welded angles, secured to top and reinforced with 1.5mm thick S.S. 25 mm x 100 mm x 25 mm channels.
4. SINK : One no. of 1.5mm thick Polished S.S. fully coved integral with work top, 16" x 16" x 8" Deep. Supplied complete with 50 mm dia waste and deck mounted faucet.
5. LEGS: Six no. of legs and all legs to be constructed of not less than 40 mm x 40 mm square 1.5mm thick S.S. Stainless steel pipe. All legs shall be uniform in finish. All legs shall be provided with stainless steel bullet feet, having an integrally formed shaft, with a minimum adjustment of approximately 40-mm without using of threading or adjusting bolts. Feet shall be completely sealed at bottom.
6. CROSSBRACING: All leg cross bracing, where required are to be constructed of not less than 25 x 25 mm square 1.2 S.S. Pipe. All cross bracing to run horizontal between all legs, approximately 150 mm above floor, unless otherwise specified. All joints to be completely welded around entire perimeter, forming complete seal, with all weld ground and painted.
7. S. S. wherever used, shall be 304 grade. Relative documents shall be submitted in technical bid.
8. Guarantee: one year from the date of successful installation at consignee place. The firm shall install this item at free of cost.

2/2

2. Revised Minimum average gross annual turnover for last three financial years & required Bid security

(NIB No. 207)

S. N.	Name of Equipment & Instruments	Indicative Quantity (In no.)	Minimum average gross annual turnover for last three financial years (In Rs.)	Minimum average gross annual turnover for last three financial years (In Rs.) for (MSMEs of Rajasthan)	Bid Security (B.S) (In Rs.)	Bid Security (B.S) for MSME Units of Rajasthan (In Rs.)
1	TABLE POST-MARTUM (Mortuaries)	200	3.00 Cr	1.50 Cr.	200000	100000

Note:-


- The above estimated quantities are only indicative and may vary substantially the above estimated quantities are only indicative for executing rate contract for a specified period. If the procuring entity does not procure any subject matter of procurement or procures less than the quantity indicated in the bidding documents the bidder shall not be entitled for any claim or compensation. No minimum quantity is guaranteed. Quantity/Capacity commitment of the firm in BF-V and clause-2.6 (xii) shall be considered for placement of supply orders.
 - Firms which are registered as micro or MSME of Rajasthan with Commissioner of Industries shall furnish the amount of bid security at the rate 0.50% of likely value of the indicative quantity or Rs. 100000/- (Rs. One Lac), whichever is less, for whole bid catalogue/each item (as per above Table-1). In respect of items for which they are registered to manufacture, shall submit an attested copy of acknowledgment of EM-II issued by DIC / Factory Licence /Udhyog Adhar, with an affidavit on non-judicial stamp paper worth Rs. 10/- as per BF-X.
- 3. In Reference to above cited subject and NIB-207, the last date for sale of bid, receipt of bid and date of opening of technical bid hereby extended as below:-**

Extended dates		
Last date for Sale of Bid form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
16.2.17 11.00 AM	16.2.17 01.00 PM	16.2.17 03.00 PM

Please note that all clarification/amendment/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms & conditions remains the same.

This bears the approval of M.D., RMSCL, Jaipur.


Executive Director (EPM)
RMSCL, Jaipur