

**Rajasthan Medical Services Corporation Limited, Jaipur**

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmsc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website: http://rmsc.health.rajasthan.gov.in

No. F-8( ) RMSC/EPM/M-2/18-19/NIB-387/ 4485

Dated: 14-12-18

**CLARIFICATION/CORRIGENDUM/ADDENDUM****Sub:- Revised Technical Specifications and bid submission date of item *Electro Surgical Unit* under NIB No. F-8( ) RMSC /EPM/M-2/18-19/NIB-387/4000 dated 27.09.18**

In Reference to subject cited above and NIB-387 the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee.

**(a) Amended Technical Specifications of item Electro Surgical Unit**

1. Microcontroller/Microprocessor/Solid State based Digital Electrosurgical Unit 300-400 Watt with digital Display Push Switch Control provides Consistent Performance for all general Surgical Procedures & Applications.
2. Unipolar as well as Bipolar facility having operating frequency between 350-700 KHz.
3. Independent Mono-polar & Bipolar Output can be used without any switch over from the machine.
4. Unit should have auto-self test facility.
5. Various Modes in Electrosurgical Cautery for Various Surgical Applications:
  - a) Mono-polar Cut:
    1. Pure cut : 200-400 W
    2. Low cut : 200-400 W
    3. Blend : 200-300 W
  - b) Mono-polar Coagulation :
    1. Spray : 120 W
    2. Fulgurate : 120 W
    3. Desiccate : 120W
  - c) Bipolar Coagulation-within :
    1. Cut : 70 -120 W
    2. Force : 70 -100 W
    3. Micro : 70 -100 W
6. All Standard Accessories to be provided(Should be of same original equipment manufacturer)-
  1. Dual pad Silicon Patient plate with cable -(Reusable- 2 Nos / Disposable-20 Nos)
  2. Bipolar forceps with Cable(Titanium non Sticky) - 2 Nos.
  3. Mono-polar hand switch pencil - (Reusable - 2 Nos / Disposable-20 Nos)
  4. Foot switch for Mono-polar and Bipolar application - 1 no.
  5. Power Cord - 1 No.
7. Standards & Safety :
  1. Product should be European CE (With Notified Body) / US FDA certified (Valid documentation should be submitted in technical bid).
  2. The unit should comply with safety requirement of IEC 60601-1 & IEC 60601-2-2 (test report should be submitted in technical bid)
8. Guarantee: Three years on equipment from the date of installation.
9. CMC: CMC shall be given @ 5% of net rate (excluding GST as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years, if required.
10. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue and preventive & calibration schedule authorized by the manufacture in the Technical bid.
11. Price Should be quoted inclusive of all accessories.
12. Training to operate the equipment & Installation will be done by supplier free of cost.
13. Service provider should be based in Rajasthan.


14. Preventive maintenance & calibration services annually during guarantee and CMC period are essential from as per the guideline of principle manufacture.
15. Demonstration of equipment is must for final technical approval, Quoted item to be demonstrated in front of technical committee should have permanent label of make, model & sr.no

In Reference to subject cited above related to NIB-387, the various representations received from the firms and issues raised by the prospective Bidders were examined by the technical committee and therefore by Purchase Committee. The following are Clarification/Corrigendum/Addendum issued for inclusion in Bid document: -

**(b) The bid is re-scheduled as follows:**

Existing Dates			Extended Dates		
Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid	Last Date for Sale of Bid Form	Last Date of Receipt of Bid Form	Date of Opening of Technical Bid
1	2	3	4	5	6
<u>17.12.2018</u> <u>11:00 AM</u>	<u>17.12.2018</u> <u>6:00 PM</u>	<u>18.12.2018</u> <u>3:00 PM</u>	<u>26.12.2018</u> <u>11:00 AM</u>	<u>26.12.2018</u> <u>6:00 PM</u>	<u>27.12.2018</u> <u>3:00 PM</u>

This corrigendum shall be signed and annexed with bid document and other terms, condition will remain same. This bears an approval of Managing Director, Rajasthan Medical Services Corporation Limited, Jaipur.

  
14.12.2018  
Executive Director (EPM)  
RMSCL, Jaipur