



## Rajasthan Medical Services Corporation Ltd.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)

Ph. No. 0141-2223887, Fax No. 0141-2228065 CIN:U24232RJ2011SGC035067 E-Mail -rmsc@nic.in, edepmrmsc-ri@nic.in

F-8(298)RMSC/EPM/M-3/2017-18/NIB-258/ 927

Date: 28/2/18

### Office Order

List of bidders declared responsive/non responsive as per documentary evaluation & technical evaluation of F-8(298)RMSC/EPM/M-3/2017-18/NIB-258/5474 Dated: 10.11.2017, technical bid opening date 04.01.2018. The list of bidders declared responsive/non responsive as per decision taken in the purchase committee dated 28.2.2018 headed by M.D., RMSCL.

S. No.	Name of Tenderer Firm	Name of Quoted Item	Responsive/ Non responsive	Remark
1.	M/s Surgicoim Medequip Pvt. Ltd., 1703-04, HSSIIDC Industrial Estate, Rai, Distt.- Sonapat, Haryana-131029 Email:-surgicoim@vsnl.com	O.T. Table (Electric)	Non Responsive	Non-Responsive Demo के दौरान फर्म द्वारा दिखाये गये इक्यूपमेंट पर मेक एंड मॉडल मेशन नहीं था।
		Examination Light	Non Responsive	Non- Responsive RMSCL specification point 9 - power consumption 15 w or less जबकि फर्म द्वारा दिखाये गये डेमो इक्यूपमेंट का power consumption 50 w था।
2.	M/s Hardik Medi-Tech, Plot No. 92, First Floor, Sector-57, Phase-IV HSIIDC, Kundli, Sonapat, Haryana Email:-anand_medical@yahoo.com	O.T. Table (Electric) and Examination Light	Responsive	Responsive
3.	M/s Confident Dental Equipments Pvt. Ltd., Plot no. 17-H, Sector I, Phase II, Bidadi Industrial Area, Bidadi-562109 Ramnagar Tq. & Dist., Karnataka State Email:-info@confidental.org	O.T. Table (Electric)	Non Responsive	Non-Responsive Point 4 should have enhanced weight bearing casters fitted with ball bearing था। तथा फर्म द्वारा weight bearing casters not fitted with ball bearing दिखाया गया।
		Examination Light	Non Responsive	Non- Responsive RMSCL specification point 9 - power consumption 15 w or less जबकि फर्म द्वारा दिखाये गये डेमो इक्यूपमेंट का power consumption 33 w था।
4.	M/s United Surgical Industries, 445, F.I.E. Patparganj Industrial Area, Delhi-110092 (India) Email:-info@unisurg.com	O.T. Table (Electric) and Examination Light	Responsive	Responsive
5.	M/s A.K. Surgical, 25/11, 1st Floor, Kaveri Path, New Sangar Road, Mansarovar, Jaipur-302 020 Email:-aksurgical@yahoo.co.in	Examination Light	Non- Responsive	Non - Responsive RMSCL specification point 9 - power consumption 15 w or less जबकि फर्म द्वारा दिखाये गये डेमो इक्यूपमेंट का power consumption 29 w था।

Online financial bids of responsive bidders for above item will be opened on dated 06.03.2018 at 11.00 A.M. Demonstration report can be obtained from office during office time.

Executive Director (EPM)  
Rajasthan Medical Services Corporation Ltd.  
Jaipur