

RAJASTHAN MEDICAL SERVICES CORPORATION LTD.
D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail - edepmrmisc-ri@nic.in

No. F-8() RMSC/EPM/M-1/17-18/NIB-299/ 2018/ 907

Dated: 28/2/18

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: - Revised Bid Schedule of item SEMI AUTOMATED URINE TEST STRIP ANALYZER and Amended Technical Specifications of Bid document of item, under No. F-8() RMSC/EPM/M-1/17-18/NIB-299/ 2017/183 date: 15.01.18

In reference to subject cited above and NIB-299, the representations received from the various firms and issues raised by the Bidders, are examined by the competent Authorities and technical committee and approved by purchase committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in Bid document Amended Technical Specification of items as below:-

1. Amended Technical Specifications of Items SEMI AUTOMATED URINE TEST STRIP ANALYZER as below:-

(01) Technical Specification of SEMI AUTOMATED URINE TEST STRIP ANALYZER

1. Analyzer meant for quantitative analysis of Urine analysts using reflectance photometry principle.
2. The analyzer must have the flexibility to select 3 different analysis modes.(General, one by one, quick mode). ~~Quick Mode is must and quick mode time should be 20-30 second~~
3. Throughput of minimum 500 tests/hour and can be upgraded to maximum.
4. Should be able to analyze – 10 parameters – Leucocytes, Nitrite, Urobilinogen, Protein, pH, Blood, Specific - Gravity, Ketones, Bilirubin, Glucose.
5. Test wave length should be 460nm to 660 nm.
6. Should have powerful sensor to automatically detect the presence of Urine Strips on the measuring area.
7. Should have touch screen/key pad for patient name and sample ID management system.
8. Should have built in higher speed thermal printer.
9. Can be connected with external bar code reader.
10. Should have storage capacity 1000-1500 patients' result and 50 QC result.
11. Should have flagging facility to demonstrate normal and abnormal results.
12. Firm should quote rates of Multi parameter strip and Albumin and Sugar strip and provide 10x100 strips of multi parameter.
13. Unit should ~~be European CE/USFDA approved.~~
14. UPS of with minimum half an hour backup should be provided.
15. **Guarantee:** Two years on equipment from the date of installation.
16. **CMC:** CMC will be given @ 5 % of Net Rate exclusive of GST (As applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for Five years (on yearly basis) after completion of Guarantee period of two years.
17. Installation will be done by supplier free of cost.
18. The service engineer should be based in Rajasthan.
19. Demonstration of equipment is must for technical approval.
20. The company should mention the make & model name of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along

with compliance sheet as per amended technical specifications.

21. Firm should mention all the pre-installation requirements in technical bid like air conditioning, dust proofing, pest proof, electrical supply with proper earthing, water supply and drainage facility.

2. **Multiparameter strips, albumin and sugar strips** की दरें सभी Bidders Bid Security जमा कराते समय पृथक से सीलड लिफाफें में अनिवार्यत प्रस्तुत करेंगे।

3. **Revised bid schedule:-**

E-bids are invited as per following revised time schedule:-

Last date & time for sale of bid form	Last date & time of receipt of bid form	Date & time of opening of technical bid
<u>14.03.18</u> 01.00 PM	<u>14.03.18</u> 06.00 PM	<u>15.03.18</u> 03.30 PM

It is also clarified that information of award of contract shall be communicated to all participating bidders on the website www.rmsc.nic.in and sppp.raj.nic.in. Please note that individual bidder will not be intimated."

Note:-

Please note that all above amendments/corrigendum in technical specifications/bid-conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

This bears the approval of M.D., RMSCL, Jaipur.


Executive Director (EPM)
RMSCL, Jaipur

