

**TECHNICAL SPECIFICATION OF PPIUCD FORCEPS**

<b>S. No.</b>	<b>Name of Item</b>	<b>Qty. in a set</b>
1	Placental forcep 315 mm	1 No.
2	Sponge Holding Forceps	1 No.
<b><u>General Terms &amp; Conditions:</u></b>		
3	All surgical instruments should be CE certified (valid documentation should be submitted in technical Bid)	
4	All surgical instruments should be manufactured by ISO 9001:2008 & ISO 13485:2003 (valid documentation should be submitted in technical Bid)	
5	Manufacturer should have ISO 7153 certificate. (valid documentation should be submitted in technical bid)	
6	Surgical instruments made by steel grade AISI 410, test report should be submitted.	
7	Laser marking along with code should be on the instrument.	
8	Guarantee: One year against any manufacturing defect.	
9	The company should mention the make & model of the quoted items and submit the technical compliance sheet in technical Bid	
10	Final technical approval after sample demonstration.	



**Executive Director (EPM)  
Rajasthan Medical Services Corporation Ltd.,  
Jaipur**