

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmsc-ri@nic.in

CIN:U24232RJ2011SGC035067

Website:www.rmsc.nic.in

No. F-8(RMSC/EPM/M-2/2016-17/NIB-196/ 46

Dated: 6/11/17.

Rate Contract and Price details

S.N.	Item	Detail
1.	Name of Equipment / Item	LED Phototherapy Unit
2.	Name & Address of Firm	M/s Fanem Medical Devices India Pvt. Ltd. ,279, 4th main, 4th phase, Ppeenya indl area, Bangalore-560058 E- mail: info@fanem.in
3.	Brand/ Model / Make	Bilitron sky 2006 make FANEM MEDICAL DEVICES INDIA PVT. LTD
4.	Packing Unit	One
5.	R/C No. & Validity	R.C. No. 352 (Valid up to 31.07.2018)
6.	Net Rate Per packing Unit (Rs.)	Rs 26,786.00 (In Words, Rs. Twenty Six Thousand Seven Hundred Eighty Six Only) + 2% concessional CST against C-form
7.	Taxes Concessional CST @ 2% (Rs.)	Rs. 535.72
8.	RMSC Surcharges @ 5%	Rs. 1,366.09
9.	Total Price (Rs.)	Rs. 28,687.81
10.	Guarantee Period	2 Years
11.	CMC after Guarantee Period	5 Years
12.	<u>Specifications of item LED Phototherapy Unit</u>	
	<ol style="list-style-type: none"> 1. Lifetime of LED 20,000 to 30,000 hours 2. Spectral Irradiance from 30 to 60 $\mu\text{W}\cdot\text{cm}^{-2}\cdot\text{nm}^{-1}$ at 45 cm distance between bed and light 3. Should have multilevel intensity control from minimum to maximum 4. At the tilted position, the irradiance is 30 $\mu\text{W}\cdot\text{cm}^{-2}\cdot\text{nm}^{-1}$ at 45 cm distance between bed and light 5. Wavelength is 450 – 480 nm, and it's free from UV and IR radiation. 6. Effective surface area - 200 X 400 mm within a irradiance ratio of 0.4 (min/max irradiance) 7. Digital (LCD) Timer for monitoring therapy hours (resettable) & lamp usage hours (non resettable) 8. Have feather touch key pad 9. Cooling Fan/ heat sink is provided to dissipate the heat created by LED's 10. Light head should be compact to use along with the Radiant warmer & provided with tilting facility so that the unit is not coming directly under warmer. 11. Smooth Height adjustment mechanism & seamless height adjustment from minimum to maximum height 12. Minimum height – 1150 mm or less from the floor to use near the mother bed 13. Maximum height – 1500 mm or more from the floor to use with the incubator and Radiant warmer. 14. Coating: Epoxy/powder coated body for scratch and rust prevention and 15. PU (Poly Urethane)/ABS/Fiber coating for plastic 16. Mobility: Three/four castors; two rear castors provided with brakes 17. The base of the unit is such that it will go beneath any Incubator/bed/trolley, with 	



	minimum of 100 mm floor clearance 18. Should be supplied with white LED for patient examination 19. The unit should be supplied with reusable eye masks, Quantity: 10 20. The product should be CE marked and the manufacturer should be ISO: 13485 certified (Certificate to be submitted) 21. The product should have regulatory approval IEC 60601-1(certificate to be submitted) 22. Two years guarantee after installation 23. CMC: CMC will be given @ 5 % (of net rate- inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after guarantee period of two years.					
13.	Comprehensive Maintenance Contract Charges (per year/per machine)					
	Rates (Rs.) Per Unit	First Year	Second Year	Third Year	Fourth Year	Fifth Year
	Comprehensive Maintenance Agreement (CMA)	CMC: CMC will be given @ 5 % (of net rate- inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after guarantee period of two years.				

RM

Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd.,

Jaipur

2