

TECHNICAL SPECIFICATION
DRESSING TROLLEY

1. Dressing Trolley should be made up of stainless steel.
2. The overall dimension of dressing trolley: 1000mm (l) x 500mm (w) x 900mm (h) ± 5 mm.
3. The dressing trolley's tabular frame should be made up of Stainless steel with 20mm ± 2 mm. dia.
4. The trolley should be with two heavy duty stainless steel shelves with dimension 750mm (l) x 500mm (w) ± 5 mm.
5. Both S.S. shelves with protective stainless steel railings on all four sides on the top.
6. Trolley should have 125mm diameter 4 castors with synthetic body, two with brake and two without brake.
7. It should have provision of S.S. Bowl and S.S. Bucket.
8. Supplied with:
 - a. Stainless steel bowl
 - b. Stainless steel bucket

Other terms and conditions:

1. Manufacturer should be ISO- 9001 certified.
2. Guarantee: should be one year for all items from the date of successfully installation at consignees address.
3. Final Technical approval after demonstration.
4. Installation will be done by supplier free of cost.
5. The Company should submit technical compliance sheet as per technical specifications mentioning the make and model of quoted item along with catalogue in technical bid.


Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd,
Jaipur