

TECHNICAL SPECIFICATIONS
SARVANGA BASHPA SWEDAN YANTHRA (LYING TYPE)

Medicinal Wood base with FRP/SS chamber and dual Boiler Chamber:-

- Base frame made of medicinal wood such as Azaderacta indica, Pterocarpus Marsupium, Albizia lebbak etc.
- Chamber must have heat resistant, wrinkled, transparent FRP body with stainless steel frame work.
- Bed made of Teakwood or PTA-FRP of high heat resistance.
- FRP condensate tray should be provided below the bed and should contain stainless steel (SS 304 Grade) steam distributors.
- Bottom portion should be adequately sealed to avoid steam leakage.
- Semi Circular collapsible head rest of at least 20 cm size to be provided.

Steam Boiler:-

- Dual heating (gas & electric) Semi-automatic boiler with water level sensing system and warning buzzer.
- Made of stainless steel (SS 304 grade) and should be provided with SS herbal basket (for herbal steam), Baumer pressure gauge, safety valve with SS spring, SS droplet remover, water level check valve, 3KW ISI quality heaters with screwed terminals, Neoprene gasket and imported Thermostat for safety.
- Provision of SS control box with a MCB, Relay system and heater indicator.
- SS flame guard (SS 202 grade), gas connection fittings, high speed burner and stand should also be provided.
- Stainless Steel Stand (SS 202 grade) for placing the boiler. Gas control valve should be anchored to the stand. Stand should have 26 cm dia. outer ring and 20 cm dia. inner ring.
- Capacity of at least 3 steaming with one filling of water.
A well balanced trolley shall be provided with the boiler.

Delivery Period - 60 Days

At



Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd,
Jaipur