



Rajasthan Medical Services Corporation Limited

Gandhi Block, SwasthyaBhawan, Tilak Marg, C-Scheme, Jaipur-302005 (Raj.)

Phone No: 0141-2228066 ,2228064

Website: <http://rmsc.health.rajasthan.gov.in>

CIN:U24232RJ2011SGC035067

E_mail :edprmasc@gmail.com

Ref. No.:- F.02(71)/RMSCL/Proc./S&S(MD)/NIB-3/2020 / 78

Dated:- 02/06/2020

Corrigendum-IV

Subject: - Date Extension and Amendments in tender quantity and technical specifications in bid and BOQ.

Ref.:- NIB No. F.02(71)/RMSCL/Proc./S&S(MD)/NIB-3/2020/35 Dated:06.03.2020

(Technical bid opening due on dated -11.06.2020)

The following amendments are made:

S. No	Existing date / Tender quantity / technical specification	Amended date / Tender quantity / technical specification												
	F.02(71)/RMSCL/Proc./S&S(MD)/NIB-3/2020/35 Dated:06.03.2020	F.02(71)/RMSCL/Proc./S&S(MD)/NIB-3/2020/35 Dated:06.03.2020												
	<table border="1"> <tr> <td>Last date and time of submission of online bids</td> <td>10.06.2020 at 6.00 PM</td> </tr> <tr> <td>EMD, Tender fees, RISL fees through challan and Physically</td> <td>10.06.2020 at 6.00 PM</td> </tr> <tr> <td>Date and time of opening of Online technical bids</td> <td>11.06.2020 at 11.00 AM</td> </tr> </table>	Last date and time of submission of online bids	10.06.2020 at 6.00 PM	EMD, Tender fees, RISL fees through challan and Physically	10.06.2020 at 6.00 PM	Date and time of opening of Online technical bids	11.06.2020 at 11.00 AM	<table border="1"> <tr> <td>Last date and time of submission of online bids</td> <td>16.06.2020 at 6.00 PM</td> </tr> <tr> <td>EMD, Tender fees, RISL fees through challan and Physically</td> <td>16.06.2020 at 6.00 PM</td> </tr> <tr> <td>Date and time of opening of Online technical bids</td> <td>17.06.2020 at 11.00 AM</td> </tr> </table>	Last date and time of submission of online bids	16.06.2020 at 6.00 PM	EMD, Tender fees, RISL fees through challan and Physically	16.06.2020 at 6.00 PM	Date and time of opening of Online technical bids	17.06.2020 at 11.00 AM
Last date and time of submission of online bids	10.06.2020 at 6.00 PM													
EMD, Tender fees, RISL fees through challan and Physically	10.06.2020 at 6.00 PM													
Date and time of opening of Online technical bids	11.06.2020 at 11.00 AM													
Last date and time of submission of online bids	16.06.2020 at 6.00 PM													
EMD, Tender fees, RISL fees through challan and Physically	16.06.2020 at 6.00 PM													
Date and time of opening of Online technical bids	17.06.2020 at 11.00 AM													

1 Annexure - VIII (List of Surgical items with Specifications) and BoQ

Annexure - VIII (List of Surgical items with Specifications) and BoQ

S. No.	Code No.	Tender Quantity
1	S-104	50000
2	S-106	800000
3	S-107	25000
4	S-108	25000
5	S-109	10000
6	S-110	10000
7	S-111	5000
8	S-112	5000
9	S-113	5000
10	S-114	5000
11	S-115	5000
12	S-116	5000
13	S-117	5000
14	S-118	5000
15	S-119	300000
16	S-120	10000

S. No.	Code No.	Tender Quantity
1	S-104	1631220
2	S-106	1323574
3	S-107	449472
4	S-108	147040
5	S-109	393556
6	S-110	52320
7	S-111	6710
8	S-112	8710
9	S-113	7680
10	S-114	6588
11	S-115	4750
12	S-116	4070
13	S-117	3510
14	S-118	2830
15	S-119	351830
16	S-120	196730

17	S-121	10000
18	S-122	10000
19	S-123	10000
20	S-124	10000
21	S-125	10000
22	S-126	5000
23	S-127	200000
24	S-128	1500000
27	S-136	1000
28	S-137	100

17	S-121	52500
18	S-122	73120
19	S-123	122050
20	S-124	39876
21	S-125	45770
22	S-126	70796
23	S-127	207108
24	S-128	439840
27	S-136	32240
28	S-137	11202

Annexure-VIII and BOQ
Technical Specification of S-110, S-120, S-121, S-122, S-123, S-127, S-137

Item code	Technical specifications	Minimum labelled Shelf life (In months)	Class As per Medical Device Rules 2017	Packing Unit
S-110	Epidural Minipack 18G, 90 mm, Metal Stylet For Single Use only, •Sterile, •Epidural catheter 20G, L 90 cm, •6 Lateral Holes closed end, •Borst Adapter •Epidural Catheter Flat filter 0.2 micro meter •Thread Assist Guide •LOR (Loss of resistance) Plastic Syringe 6 ml	36 month	Class B	Each Piece
S-120	3 Way Stop Cock with extension Tube (Vein -O-Line) size 10 cm (Non Pyrogenic & Single Use)	36 month	Class B	Each Piece
S-121	3 Way Stop Cock with extension Tube (Vein -O-Line) size 50 cm (Non Pyrogenic & Single Use)	36 month	Class B	Each Piece
S-122	3 Way Stop Cock with extension Tube (Vein -O-Line) size 100 cm (Non Pyrogenic &	36 month	Class B	Each Piece

Annexure-VIII and BOQ are to be read as follows:-
Technical Specification of S-110, S-120, S-121, S-122, S-123, S-127, S-137

Item code	Technical specifications	Minimum labelled Shelf life (In months)	Class As per Medical Device Rules 2017	Packing Unit
S-110	Epidural Minipack 18G, 80mm - 90mm , Metal Stylet For Single Use only, •Sterile, •Epidural catheter 20G, L 90 cm, • 3 to 6 Lateral Holes closed end, •Borst Adapter •Epidural Catheter Flat filter 0.2 micro meter •Thread Assist Guide •LOR (Loss of resistance) • plastic syringe 6ml/10ml	36 month	Class B	Each Piece
S-120	3 Way Stop Cock with extension Tube (Vein-O Extension line) size 10 cm (Non Pyrogenic & Single Use)	36 month	Class B	Each Piece
S-121	3 Way Stop Cock with extension Tube (Vein-O Extension line) size 50 cm (Non Pyrogenic & Single Use)	36 month	Class B	Each Piece
S-122	3 Way Stop Cock with extension Tube (Vein-O	36 month	Class B	Each Piece



Single Use)									
S-123	3 Way Stop Cock with extension Tube (Vein -O-Line) size 150 cm (Non Pyrogenic & Single Use)	36 month	Class B	Each Piece	Extension line) size 100 cm (Non Pyrogenic & Single Use)				
S-127	Elastic Adhesive Bandage BP 1m x 10cm Adhesive material should have good quality sticking property, non-allergic	36 month	Class B	Each Piece	S-123	3 Way Stop Cock with extension Tube (Vein-O Extension line size 150 cm (Non Pyrogenic & Single Use)	36 month	Class B	Each Piece
S-137	Chemotherapy Port & Non-coring needles (Pediatric) Valved catheter need only saline flush, catheter with intermediate size power port with large septum and silicon filled suture holes, should be MRI compatible with cathlock radio-opaque ring. 6FR with silicon material with peel apart percutaneous introducer system Chemo port huber needle 20g and 22g.	36 month	Class B	Each Piece	S-127	Elastic Adhesive Bandage BP 10 cm x 1 mtr Stretched Length Adhesive material should have good quality sticking property, non-allergic	36 month	Class B	Each Piece
					S-137	Chemotherapy Port & Non-coring needles (Pediatric) catheter need only saline flush, catheter with intermediate size power port with large septum and silicon filled suture holes, should be MRI compatible with cathlock radio-opaque ring. 6FR with silicon material with peel apart percutaneous introducer system Chemo port huber needle 20g and 22g.	36 month	Class B	Each Piece

Note:-

- Rate should be quoted in BoQ as per amended technical specifications.
- It may be noted that if any further amendments are issued then a corrigendum will be published and informed.
- Rest of the terms and conditions will remain the same.

[Signature]
 Executive Director (Proc.)
 RMSCL

