



Government of Rajasthan
Medical & Health Department

Rajasthan Medical Services Corporation, Swasthaya Bhawan, C-Scheme, Jaipur

Phone No: 0141-2228065, Fax No: 0141-2228065

CIN: U24232RJ2011SGC035067

E_mail : rmisc@nic.in

Website: WWW.rmisc.nic.in

Ref. F.02()/RMSC/PROCUREMENT/S&S/NIB-02/2016/ 77

Dated: 17.03.16

Corrigendum-II


Subject:- Amended conditions of bid document for the NIB No. F.02(22) /RMSC /PROCUREMENT/S&S/NIB-02/2016/37 Dated:09.02.2016 (For Drug Items)

The following amendments are made:

1. In Bid document's clause 2(b), 2(c), 5(m), 5(n), Annexure III (Annual Turnover statement) and Annexure XV (Performance statement) financial year 2012-13, 2013-14, 2014-15 are mentioned, it is amended as:-
2012-13, 2013-14, 2014-15 or 2013-14, 2014-15, 2015-16
2. Relaxation Regarding Market Standing Certificate [Clause 2(c), 5(i)(j)]:- Two year Market Standing Certificate is also acceptable instead of three years for Item code S-15(a) and S-15(c).

It bears approval of Purchase Committee dated 14.03.2016 of RMSC.

Rest terms and conditions will remain the same.


Executive Director (Proc.)
RMSC

Government of Rajasthan
Medical & Health Department

Rajasthan Medical Services Corporation, Swasthaya Bhawan, C-Scheme, Jaipur

Phone No: 0141-2228065, Fax No: 0141-2228065

E mail : rmsc@nic.in

CIN: U24232RJ2011SGC035067

Website: WWW.rmhc.nic.in

Ref. F.02()/RMSC/PROCUREMENT/S&S/NIB-02/2016/ 79

Dated: 17.03.2016

Corrigendum-III

Subject:- Amended specification of bid document for the NIB No. F.02(22) /RMSC /PROCUREMENT/S&S/NIB-02/2016/37 Dated:09.02.2016 (For Drug Items)

The following amendments are made; in Annexure XIV and BOA

S. No	Item Code	Existing specification	Amended specification
1	S-4	Blood Administration Set / Blood Transfusion Set <ul style="list-style-type: none"> Sharp and easy piercing spike suitable for blood bags and standard blood containers Transparent cylindrical drip chamber with 200 micron filter 150 cm long smooth kink resistant tubing Efficient roller clamp to control and adjust the transfusion rate Should conform to ISO 1135-4 standard	Blood Administration Set / Blood Transfusion Set <ul style="list-style-type: none"> Sharp and easy piercing spike suitable for blood bags and standard blood containers Transparent cylindrical drip chamber with filter. Filter Size should be 200+20 micrometer. 150 cm long smooth kink resistant tubing Efficient roller clamp to control and adjust the transfusion rate Should conform to IS 9824 (Part 3) :1996
2	S-8(d)	Suction Catheter, Sterile. Size: FG 10 <ul style="list-style-type: none"> Soft, kink resistant tubing Rounded open tip with lateral eye Colour coded universal funnel connector for safe connection to standard suction equipment Length 50 cm (min.)	Suction Catheter, Sterile. Size: FG 10 (for use in respiratory tract) <ul style="list-style-type: none"> Soft, kink resistant tubing Rounded open tip with lateral eye Colour coded universal funnel connector for safe connection to standard suction equipment Length 50 cm (min.) Should conform ISO 8836 : 2014
3	S-21	Plaster of Paris Bandages 15cm X 2.7mts / Roll	<ul style="list-style-type: none"> Plaster of Paris Bandages 15cm X 2.7mts / Roll Should conform to Schedule F(II) of Drug & Cosmetic Act 1940
4	S-22	Plaster of Paris Bandages 10cm X 2.7mts / Roll	<ul style="list-style-type: none"> Plaster of Paris Bandages 10cm X 2.7mts / Roll Should conform to Schedule F(II) of Drug & Cosmetic Act 1940

(Handwritten signatures and initials)

मुख्यमंत्री निःशुल्क दवा योजना



S. No	Item Code	Existing specification	Amended specification
5	S-79	Sterilized Umbilical Cotton Tape Width 3 mm, Length 75 cm	<ul style="list-style-type: none"> • Sterilized Umbilical Cotton Tape Width 3 mm, Length 75 cm • Should conform to Schedule F(III) of Drug & Cosmetic Act 1940
6	R-41	NON ABSORBABLE SURGICAL SUTURE, STERILISED SURGICAL NEEDLED SUTURE MONOFILAMENT POLYPROPYLENE BLUE Polypropylene (3/8 Cir Conventional Cutting PC-3Needle 15mm Length 60cm)	NON ABSORBABLE SURGICAL SUTURE, STERILISED SURGICAL NEEDLED SUTURE USP MONOFILAMENT POLYPROPYLENE BLUE Polypropylene (3/8 Cir Conventional Cutting PC-3Needle 15mm Length 60cm)

It bears approval of Purchase Committee dated 14.03.2016 of RMSC.

Rest terms and conditions will remain the same.

[Handwritten mark]

[Handwritten mark]

[Handwritten signature]
Executive Director (Proc.)
RMSC