

**Rajasthan Medical Services Corporation, Swasthaya Bhawan, C-Scheme, Jaipur**

Phone No: 0141-2228065, Fax No: 0141-2228065

CIN:U24232RJ2011SGC035067

Email: [rmsc@nic.in](mailto:rmsc@nic.in)

[www.rmsc.health.rajasthan.gov.in](http://www.rmsc.health.rajasthan.gov.in)

Ref No F.02(40)/RMSC/Proc./Surgical (Non Drug)/NIB-01/2017/ 169

Dated: 25/04/2017

**Tentative List of Bidders Declared Responsive/Non Responsive**

(Tender No F.02(40)/RMSC/Proc./Surgical (Non Drug)/NIB-01/2017/76 Dated: 17.02.2017

Technical bid Opened on 22.03.2017)

The Details of Bidders Declared Responsive/Non Responsive as decided in the Purchase Committee meeting dated on 25.04.2017 headed by MD RMSCL.

S. No	Name of Bidder	Quoted Item No.	Responsive Items code no.	Non Responsive Items with reasons
1	M/s Anondita Healthcare	S-5(a), S-5(b), S-6(a), S-6(b), S-7(a), S-7(b)	S-5(a), S-6(a), S-7(a)	S-5(b), S-6(b), S-7(b) Not Approved by technical/ clinical evaluation committee because rubber quality is not good.
2	M/s Cadila Pharamaceutical Ltd.	S-5(a), S-5(b), S-6(a), S-6(b), S-7(a), S-7(b)	Nil	S-5(a), S-5(b), S-6(a), S-6(b), S-7(a), S-7(b) Due to 1. Performance statement submitted by firm does not fulfill 15% of tender quantity requirement. 2. Production capacity less than supply commitment quantity 3.S-5(b), S-6(b), S-7(b) Not Approved by technical/ clinical evaluation committee because rubber quality is not good.
3	M/s Hi-Care Polymers Pvt. Ltd.	S-5(a), S-5(b), S-6(a), S-6(b), S-7(a), S-7(b)	Nil	S-5(a), S-5(b), S-6(a), S-6(b), S-7(a), S-7(b) Not Approved by technical/ clinical evaluation committee because rubber quality is not good.
4	M/s HLL Lifecare Ltd.	S-99(b)	Nil	1. Firm does not submit Performance statement as per tender requirement. 2. Firm has not submitted required clarification / documents.
5	M/s Jackson Care Product	S-99(b)	Nil	Firm has not submitted any relevant document about the date from which the firm has started manufacturing the Belt type Sanitary Napkin as per RMSC Specification.
6	M/s MRK Healthcare Pvt. Ltd.	S-5(a), S-5(b), S-6(a), S-6(b), S-7(a), S-7(b)	S-5(a), S-5(b), S-6(a), S-6(b), S-7(a), S-7(b)	-
7	M/s RFB Latex Ltd.	S-5(a), S-5(b), S-6(a), S-6(b), S-7(a), S-7(b)	Nil	S-5(a), S-5(b), S-6(a), S-6(b), S-7(a), S-7(b) Not Approved by technical/ clinical



मुख्यमंत्री निःशुल्क दवा योजना



S. No	Name of Bidder	Quoted Item No.	Responsive Items code no.	Non Responsive Items with reasons
				evaluation committee because rubber quality is not good.
8	M/s Safecare Rubber Products Pvt. Ltd.	S-5(a), S-5(b), S-6(a), S-6(b), S-7(a), S-7(b)	Nil	S-5(a), S-6(a), S-7(a), S-5(b), S-6(b), S-7(b) 1. Due to supply commitment quantity and production capacity is less than tender quantity 2. S-5(a), S-6(a), S-7(a) Not Approved by technical/ clinical evaluation committee because rubber quality is not good. 3. S-5(b), S-6(b), S-7(b) Sample Not submitted by the firm.
9	M/s St. Marys Rubbers Pvt. Ltd.	S-5(a), S-5(b), S-6(a), S-6(b), S-7(a), S-7(b)	Nil	S-5(a), S-6(a), S-7(a), S-5(b), S-6(b), S-7(b) Not Approved by technical/ clinical evaluation committee because rubber quality is not good.
10	M/s Sutures India Pvt. Ltd.	S-81, S-5(a), S-5(b), S-6(a), S-6(b), S-7(a), S-7(b)	S-5(a), S-5(b), S-6(a), S-6(b), S-7(a), S-7(b)	S-81 Not Approved by technical/ clinical evaluation committee because wire does not stay at place.
11	M/s Vijay Latex Products Pvt. Ltd.	S-5(a), S-6(a), S-7(a)	Nil	S-5(a), S-6(a), S-7(a) Not Approved by technical/ clinical evaluation committee due to inappropriate size.

Online Financial Bids (BoQ) of Responsive Bidders will be opened on 28.04.2017 at 3.00 P.M. The above list will be treated as final if no objection is received till 28.04.2017 at 11:00 A.M. or no change otherwise is required in the list after examination of any representation.

*(Handwritten marks)*

*(Signature)*  
Executive Director (Proc.)  
RMSC