Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax: 0141-2228065

E-mail: edepmrmsc-ri@nic.in

CIN:U24232RJ2011SGC035067

Website: www.rmsc.health.rajasthan.gov.in-

No. F-8()RMSC/EPM/M-2/2016-17/NIB-196/

Dated:

Order For Rate Change In Rate Contract Due To GST Provision					
S.N.	Item	Detail	N 3		
1.	Name of Equipment / Item	LED Phototherapy Unit			
2.	Name & Address of Firm	M/s Fanem Medical Devices India Pvt. Ltd. ,279-4th main, 4th Phase, Pecnya Ind Bangalore-560058 E- mail: info@fanem.in	ll Area,		
3.	Brand/ Model / Make	Bilitron sky 2006 make FANEM MEDICAL DEVICES INDIA PVT. LTD			
4.	Packing Unit	One	1		
5.	R/C No. & Validity	R.C. No. 352 (Valid up to 31.07.2018)			
6.	Net Rate Per packing Unit (Rs.)	Rs 26,786.00 (In Words, Rs. Twenty Six Thousand Seven Hundred Eighty Six Only) + 2% concession against C-form	nal CST		
7.	Concessional CST @ 2%	Rs. 535.72			
8.	RMSC Surcharges @ 5%	Rs. 1,366.09			
9.	Total Approved Rate Including Concessional; CST and RMSC Surcharge	Rs. 28,687.81			
10.		Basic New Rate (HSN Code 9018)	-		
	(i) Existing Approved Net Rate in	Date Control (T. D.)	26786.00		
	(ii) Less Excise Duty (In Rs.)				
	(iii) Less any other amount intimat	ed by the firm (In Rs.)	1516.00		
	(iv) New Basic Rate {(i)-(ii)-(iii)}	(In Pa)	0.00 (
11.		Revised Rates for Supply to RMSC from Suppliers	25270.00		
	(i) New Basic Rate (In Rs.)		<u>Re 1878 (1888)</u>		
	(ii) IGST @ 12 % (In Rs.)		25270.00		
	(iii) CGST (In Rs.)		3032.40		
Ī	(iv) SGST (In Rs.)		0.00		
	(v) Approved Rate For Supplier	To RMSC ((i)+(ii)+(iii)+(iv) (In Pa)	0.00		
	(v) Approved Rate For Supplier To RMSC {(i)+(ii)+(iii)+(iv) (In Rs.)				





	G Even DMSC	
12.	Revised Rate For Supply to Consignee From RMSC	25270.00
	(i) New Basic Rate (In Rs.)	1263.50
	(ii) RMSC Surcharge @ 5 % On (i) (In Rs.)	1592.01
,	(iii) CGST@ 6 % { On (i)+(ii)} (In Rs.)	1592.01
	$\alpha = \alpha =$	29717.52
	(v) Approved Rate For Consignee From RMSC {(i)+(ii)+(iii)+(iv) (ln Rs.)	

Executive Director (EPM) RMSC, Jaipur

Dated:

No. E-8()RMSC/EPM/M-2/2016-17/NYB-196/

Copy forwarded to the following for information & necessary action:

P.S. to Mission Director NHM Jaipur.

pirector, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).

Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.

Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Scinces, Jaipur).

Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.

All P.M.O.'s.

All CM&HO's.

Executive Director (Logistics/Finance)

9. AGM (IT), RMSC, Jaipur

11. M/s Fanem Medical Devices India Pvt. Ltd. ,279, 4th main, 4th Phase, Peenya Indl Area, Bangalore-560058

12. Guard file.

2002 Executive Director (EPM) RMSC, Jaipur

1.

Specifications of item LED Phototherapy Unit

- 1. Lifetime of LED 20,000 to 30,000 hours
- 2. Spectral Irradiance from 30 to 60 $\mu W \cdot cm^{-2} \cdot nm^{-1}$ at 45 cm distance between bed and light
- 3. Should have multilevel intensity control from minimum to maximum
- 4. At the tilted position, the irradiance is 30 μW·cm⁻²·nm⁻¹ at 45 cm distance between bed and light
- 5. Wavelength is 450 480 nm, and it's free from UV and IR radiation.
- 6. Effective surface area 200 X 400 mm within a irradiance ratio of 0.4 (min/max irradiance)
- 7. Digital (LCD) Timer for monitoring therapy hours (resettable) & lamp usage hours (non resettable)
- 8. Have feather touch key pad
- 9. Cooling Fan/ heat sink is provided to dissipate the heat created by LED's
- 10. Light head should be compact to use along with the Radiant warmer & provided with tilting facility so that the unit is not coming directly under warmer.
- 11. Smooth Height adjustment mechanism & seamless height adjustment from minimum to maximum height
- 12. Minimum height 1150 mm or less from the floor to use near the mother bed
- 13. Maximum height 1500 mm or more from the floor to use with the incubator and Radiant warmer.
- 14. Coating: Epoxy/powder coated body for scratch and rust prevention and
- 15. PU (Poly Urethane)/ABS/Fiber coating for plastic
- 16. Mobility: Three/four castors; two rear castors provided with brakes
- 17. The base of the unit is such that it will go beneath any Incubator/bed/trolley, with minimum of 100 mm floor clearance
- 18. Should be supplied with white LED for patient examination
- 19. The unit should be supplied with reusable eye masks, Quantity: 10
- 20. The product should be CE marked and the manufacturer should be ISO: 13485 certified (Certificate to be submitted)
- 21. The product should have regulatory approval IEC 60601-1(certificate to be submitted)
- 22. Two years guarantee after installation
- 23. CMC: CMC will be given @ 5 % (of net rate- inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after guarantee period of two years.